NBS
PUBLICATIONS

NAT'L INST. OF STAND & TECH R.I.C.

A11103 995864

NBS SPECIAL PUBLICATION 260-104

U.S. DEPARTMENT OF COMMERCE/National Bureau of Standards

Standard Reference Materials:

Summary of the Biological and Botanical Standards Issued by the National Bureau of Standards

R. Mavrodineanu and R. Alvarez

QC
100
.U57
260-104
C. 2
1985

The National Bureau of Standards[1] was established by an act of Congress on March 3, 1901. The Bureau's overall goal is to strengthen and advance the nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research and provides: (1) a basis for the nation's physical measurement system, (2) scientific and technological services for industry and government, (3) a technical basis for equity in trade, and (4) technical services to promote public safety. The Bureau's technical work is performed by the National Measurement Laboratory, the National Engineering Laboratory, the Institute for Computer Sciences and Technology, and the Institute for Materials Science and Engineering.

The National Measurement Laboratory

Provides the national system of physical and chemical measurement; coordinates the system with measurement systems of other nations and furnishes essential services leading to accurate and uniform physical and chemical measurement throughout the Nation's scientific community, industry, and commerce; provides advisory and research services to other Government agencies; conducts physical and chemical research; develops, produces, and distributes Standard Reference Materials; and provides calibration services. The Laboratory consists of the following centers:

- Basic Standards[2]
- Radiation Research
- Chemical Physics
- Analytical Chemistry

The National Engineering Laboratory

Provides technology and technical services to the public and private sectors to address national needs and to solve national problems; conducts research in engineering and applied science in support of these efforts; builds and maintains competence in the necessary disciplines required to carry out this research and technical service; develops engineering data and measurement capabilities; provides engineering measurement traceability services; develops test methods and proposes engineering standards and code changes; develops and proposes new engineering practices; and develops and improves mechanisms to transfer results of its research to the ultimate user. The Laboratory consists of the following centers:

- Applied Mathematics
- Electronics and Electrical Engineering[2]
- Manufacturing Engineering
- Building Technology
- Fire Research
- Chemical Engineering[2]

The Institute for Computer Sciences and Technology

Conducts research and provides scientific and technical services to aid Federal agencies in the selection, acquisition, application, and use of computer technology to improve effectiveness and economy in Government operations in accordance with Public Law 89-306 (40 U.S.C. 759), relevant Executive Orders, and other directives; carries out this mission by managing the Federal Information Processing Standards Program, developing Federal ADP standards guidelines, and managing Federal participation in ADP voluntary standardization activities; provides scientific and technological advisory services and assistance to Federal agencies; and provides the technical foundation for computer-related policies of the Federal Government. The Institute consists of the following centers:

- Programming Science and Technology
- Computer Systems Engineering

The Institute for Materials Science and Engineering

Conducts research and provides measurements, data, standards, reference materials, quantitative understanding and other technical information fundamental to the processing, structure, properties and performance of materials; addresses the scientific basis for new advanced materials technologies; plans research around cross-country scientific themes such as nondestructive evaluation and phase diagram development; oversees Bureau-wide technical programs in nuclear reactor radiation research and nondestructive evaluation; and broadly disseminates generic technical information resulting from its programs. The Institute consists of the following Divisions:

- Ceramics
- Fracture and Deformation[3]
- Polymers
- Metallurgy
- Reactor Radiation

[1] Headquarters and Laboratories at Gaithersburg, MD, unless otherwise noted; mailing address Gaithersburg, MD 20899.
[2] Some divisions within the center are located at Boulder, CO 80303.
[3] Located at Boulder, CO, with some elements at Gaithersburg, MD.

NBS
RESEARCH INFORMATION
CENTER

NBS

Standard Reference Materials:

Summary of the Biological and Botanical Standards Issued by the National Bureau of Standards

R. Mavrodineanu[1]

Center for Analytical Chemistry
National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

and

R. Alvarez

Office of Standard Reference Materials
National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

[1]Present Address: 227 Almeria Road
West Palm Beach, FL 33405

U.S. DEPARTMENT OF COMMERCE, Malcolm Baldrige, Secretary
NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director

Issued October 1985

Library of Congress Catalog Card Number: 85-600605

National Bureau of Standards Special Publication 260-104
Natl. Bur. Stand. (U.S.), Spec. Publ. 260-104, 68 pages (Oct. 1985)
CODEN: XNBSAV

U.S. GOVERNMENT PRINTING OFFICE
WASHINGTON: 1985

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402

Preface

Standard Reference Materials (SRM's) as defined by the National Bureau of Standards (NBS) are well-characterized materials, produced in quantity and certified for one or more physical or chemical properties. They are used to assure the accuracy and compatibility of measurements throughout the Nation. SRM's are widely used as primary standards in many diverse fields in science, industry, and technology, both within the United States and throughout the world. They are also used extensively in the fields of environmental and clinical analysis. In many applications, traceability of quality control and measurement processes to the national measurement system is carried out through the mechanism and use of SRM's. For many of the Nation's scientists and technologists it is therefore of more than passing interest to know the details of the measurements made at NBS in arriving at the certified values of the SRM's produced. An NBS series of papers, of which this publication is a member, called the NBS Special Publication - 260 Series, is reserved for this purpose.

This 260 Series is dedicated to the dissemination of information on different phases of the preparation, measurement, certification and use of NBS SRM's. In general, much more detail will be found in these papers than is generally allowed, or desirable, in scientific journal articles. This enables the user to assess the validity and accuracy of the measurement processes employed, to judge the statistical analysis, and to learn details of techniques and methods utilized for work entailing the greatest care and accuracy. These papers also should provide sufficient additional information not found on the certificate so that new applications in diverse fields not foreseen at the time the SRM was originally issued will be sought and found.

Inquiries concerning the technical content of this paper should be directed to the author(s). Other questions concered with the availability, delivery, price, and so forth, will receive prompt attention from:

Office of Standard Reference Materials
National Bureau of Standards
Gaithersburg, MD 20899

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

Catalog of NBS Standard Reference Materials (1984-85 edition), Catherine H. Hudson, ed., NBS Spec. Publ. 260 (February 1984). $5.50* SN003-003-02558-5.

Michaelis, R. E., and Wyman, L. L. Standard Reference Materials: Preparation of White Cast Iron Spectrochemical Standards. NBS Misc. Publ. 260-1 (June 1964). COM74-11061**

Michaelis, R. E., Wyman, L. L., and Flitsch, R., Standard Reference Materials: Preparation of NBS Copper-Base Spectrochemical Standards. NBS Misc. Publ. 260-2 (October 1964). COM74-11063**

Michaelis, R. E., Yakowitz, H., and Moore, G. A., Standard Reference Materials: Metallographic Characterization of an NBS Spectrometric Low-Alloy Steel Standard. NBS Misc. Publ. 260-3 (October 1964). COM74-11060**

Hague, J. L. Mears, T. W., and Michaelis, R. E., Standard Reference Materials: Sources of Information, NBS Misc. Publ. 260-4 (February 1965). COM74-11059

Alvarez, R., and Flitsch R., Standard Reference Materials: Accuracy of Solution X-Ray Spectrometric Analysis of Copper-Base Alloys. NBS Misc. Publ. 260-5 (March 1965). PB168068**

Shultz, J. I., Standard Reference Materials: Methods for the Chemical Analysis of White Cast Iron Standards, NBS Misc. Publ. 260-6 (July 1965). COM74-11068**

Bell, R. K., Standard Reference Materials: Methods for the Chemical Analysis of NBS Copper-Base Spectrochemical Standards. NBS Misc. Publ. 260-7 (October 1965). COM74-11067**

Richmond, M.S., Standard Reference Materials: Analysis of Uranium Concentrates at the National Bureau of Standards. NBS Misc. Publ. 260-8 (December 1965). COM74-11066**

Anspach, S. C., Cavallo, L. M. Garfinkel, S. B. Hutchinson, J. M. R., and Smith, C. N., Standard Reference Materials: Half Lives of Materials Used in the Preparation of Standard Reference Materials of Nineteen Radioactive Nuclides Issued by the National Bureau of Standards NBS Misc. Publ. 260-9 (November 1965). COM74-11065**

Yakowitz, H., Vieth, D. L., Heinrich, K. F. J., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards II: Cartridge Brass and Low-Alloy Steel, NBS Misc. Publ. 260-10 (December (1965). COM74-11064**

Napolitano, A., and Hawkins, E. G., Standard Reference Materials: Viscosity of Standard Lead-Silica Glass, NBS Misc. Publ. 260-11 (November 1966). NBS Misc. Publ. 260-11**

Yakowitz, H., Vieth, D. L., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards III: White Cast Iron and Stainless Steel Powder Compact, NBS Misc. Publ. 260-12 (September 1966). NBS Misc. Publ. 260-12**

Spijkerman, J. L., Snediker, D. K., Ruegg, F. C., and DeVoe, J. R., Standard Reference Materials: Mossbauer Spectroscopy Standard for the Chemical Shift of Iron Compounds, NBS Misc. Publ. 260-13 (July 1967). NBS Misc. Publ. 260-13**

Menis, O., and Sterling, J. T., Standard Reference Materials: Determination of Oxygen in Ferrous Materials - SRM 1090, 1091, and 1092, NBS Misc. Publ. 260-14 (September 1966). NBS Misc. Publ. 260-14**

Passaglia, E., and Shouse, P. J., Standard Reference Materials: Recommended Method of Use of Standard Light-Sensitive Paper for Calibrating Carbon Arcs Used in Testing Textiles for Colorfastness to Light, NBS Misc. Publ. 260-15 (June 1967). (Replaced by NBS Spec. Publ. 260-41.)

Yakowitz, H., Michaelis, R. E., and Vieth, D. L., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards IV: Preparation and Microprobe Characterization of W-20% MO Alloy Fabricated by Powder Metallurgical Methods, NBS Spec. Publ. 260-16 (January 1969). COM74-11062**

Catanzaro, E. J., Champion, C. E., Garner, E. L., Marinenko, G., Sappenfield, K. M., and Shields, W. R., Standard Reference Materials: Boric Acid; Isotopic and Assay Standard Reference Materials, NBS Spec. Publ. 260-17 (February 1970). Out of Print.

Geller, S. B., Mantek, P.A., and Cleveland, N. G., Standard Reference Materials: Calibration of NBS Secondary Standard Magnetic Tape (Computer Amplitude Reference) Using the Reference Tape Amplitude Measurement "Process A," NBS Spec. Publ. 260-18 (November 1969). (See NBS Spec. Publ. 260-29.)

Paule, R. C., and Mandel, J., Standard Reference Materials: Analysis of Interlaboratory Measurements on the Vapor Pressure of Gold (Certification of Standard Reference Material 745). NBS Spec. Publ. 260-19 (January 1970). PB190071**

Paule, R. C., and Mandel, J., Standard Reference Materials: Analysis of Interlaboratory Measurements on the Vapor Pressures of Cadmium and Silver, NBS Spec. Publ. 260-21 (January 1971). COM74-11359**

Yakowitz, H., Fiori, C. E., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of Fe-3 Si Alloy, NBS Spec. Publ. 260-22 (February 1971). COM74-11357**

Napolitano, A., and Hawkins, E. G., Standard Reference Materials: Viscosity of a Standard Borosilicate Glass, NBS Spec. Publ. 260-23 (December 1970). COM71-00157**

Sappenfield, K. M., Marineko, G., and Hague, J. L., Standard Reference Materials: Comparison of Redox Standards, NBS Spec. Publ. 260-24 (January 1972). COM72-50058**

Hicho, G. E., Yakowitz, H., Rasberry, S. D., and Michaelis, R. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Four Percent Austenite, NBS Spec. Publ. 260-25 (February 1971). COM74-11356**

Martin, J. F., Standard Reference Materials: National Bureau of Standards-US Steel Corporation Joint Program for Determining Oxygen and Nitrogen in Steel, NBS Spec. Publ. 260-26 (February 1971). 85 cents* PB 81176620

Garner, E. L., Machlan, L. A., and Shields, W. R., Standard Reference Materials: Uranium Isotopic Standard Reference Materials, NBS Spec. Publ. 260-27 (April 1971). COM74-11358**

Heinrich, K. F. J., Myklebust, R. L., Rasberry, S. D., and Michaelis, R. E., Standard Reference Materials: Preparation and Evaluation of SRM's 481 and 482 Gold-Silver and Gold-Copper Alloys for Microanalysis, NBS Spec. Publ. 260-28 (August 1971). COM71-50365**

Geller, S. B., Standard Reference Materials: Calibration of NBS Secondary Standard Magnetic Tape (Computer Amplitude Reference) Using the Reference Tape Amplitude Measurement "Process A-Model 2," NBS Spec. Publ. 260-29 (June 1971). COM71-50282

Gorozhanina, R. S., Freedman, A. Y., and Shaievitch, A. B. (translated by M. C. Selby), Standard Reference Materials: Standard Samples Issued in the USSR (A Translation from the Russian). NBS Spec. Publ. 260-30 (June 1971). COM71-50283**

Hust, J. G., and Sparks, L. L., Standard Reference Materials: Thermal Conductivity of Electrolytic Iron SRM 734 from 4 to 300 K, NBS Spec. Publ. 260-31 (November 1971). COM71-50563**

Mavrodineanu, R., and Lazar, J. W., Standard Reference Materials: Standard Quartz Cuvettes, for High Accuracy Spectrophotometry, NBS Spec. Publ. 260-32 (December 1973). 55 cents* SN003-003-01213-1

Wagner, H. L., Standard Reference Materials: Comparison of Original and Supplemental SRM 705, Narrow Molecular Weight Distribution Polystyrene, NBS Spec. Publ. 260-33 (May 1972). COM72-50526**

Sparks, L. L., and Hust, J. G., Standard Reference Materials: Thermoelectric Voltage, NBS Spec. Publ. 260-34, (April 1972). COM72-50371**

Sparks, L. L., and Hust, J. G., Standard Reference Materials: Thermal Conductivity of Austenitic Stainless Steel, SRM 735 from 5 to 280 K, NBS Spec. Publ. 260-35 (April 1972.) 35 cents* COM72-50368**

Cali, J. P., Mandel, J., Moore, L. J., and Young, D. S., Standard Reference Materials: A Referee Method for the Determination of Calcium in Serum, NBS SRM 915, NBS Spec. Publ. 260-36 (May 1972). COM72-50527**

Shultz, J. I. Bell., R. K. Rains, T. C., and Menis, O., Standard Reference Materials: Methods of Analysis of NBS Clay Standards, NBS Spec. Publ. 260-37 (June 1972). COM72-50692**

Richmond, J. C., and Hsia, J. J., Standard Reference Materials: Preparation and Calibration of Standards of Spectral Specular Reflectance, NBS Spec. Publ. 260-38 (May 1972). COM72-50528**

Clark, A. F., Denson, V.A., Hust, J. G., and Powell, R. L., Standard Reference Materials: The Eddy Current Decay Method for Resistivity Characterization of High-Purity Metals, NBS Spec. Publ. 260-39 (May 1972). COM72-50529**

McAdie, H. G., Garn, P.D., and Menis, O., Standard Reference Materials: Selection of Thermal Analysis Temperature Standards Through a Cooperative Study (SRM 758, 759, 760), NBS Spec. Publ. 260-40 (August 1972.) COM72-50776**

Wood, L. A., and Shouse, P. J., Standard Reference Materials: Use of Standard Light-Sensitive Paper for Calibrating Carbon Arcs Used in Testing Textiles for Colorfastness to Light, NBS Spec. Publ. 260-41 (August 1972) COM72-50775**

Wagner, H. L. and Verdier, P. H., eds., Standard Reference Materials: The Characterization of Linear Polyethylene, SRM 1475, NBS Spec. Publ. 260-42 (September 1972). COM72-50944**

Yakowitz, H., Ruff, A. W., and Michaelis, R. E., Standard Reference Materials: Preparation and Homogeneity Characterization of an Austenitic Iron-Chromium-Nickel Alloy, NBS Spec. Publ. 260-43 (November 1972). COM73-50760**

Schooley, J. F., Soulen, R. J., Jr., and Evans, G. A., Jr., Standard Reference Materials: Preparation and Use of Superconductive Fixed Point Devices, SRM 767, NBS Spec. Publ. 260-44 (December 1972). COM73-50037**

Greifer, B., Maienthal, E. J., Rains, T. C., and Rasberry, S. D., Standard Reference Materials: Powdered Lead-Based Paint, SRM 1579, NBS Spec. Publ. 260-45 (March 1973). COM73-50226**

Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials: Austenitic Stainless Steel, SRM's 735 and 798, from 4 to 1200 K, NBS Spec. Publ. 260-46 (March 1975). SN003-003-01278-5*

Hust, J. G., Standard Reference Materials: Electrical Resistivity of Electrolytic Iron, SRM 797, and Austenitic Stainless Steel, SRM 798, from 5 to 280 K, NBS Spec. Publ. 260-47 (February 1974). COM74-50176**

Mangum, B. W., and Wise, J. A., Standard Reference Materials: Description and Use of Precision Thermometers for the Clinical Laboratory, SRM 933 and SRM 934, NBS Spec. Publ. 260-48 (May 1974). 60 cents* SN003-003-01278-5

Carpenter, B. S., and Reimer, G. M., Standard Reference Materials: Calibrated Glass Standards for Fission Track Use, NBS Spec. Publ. 260-49 (November 1974). COM74-51185

Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials: Electrolytic Iron, SRM's 734 and 797 from 4 to 1000 K, NBS Spec. Publ. 260-50 (June 1975). $1.00* SN003-003-01425-7

Mavrodineanu, R., and Baldwin, J. R., Standard Reference Materials: Glass Filters As a Standard Reference Material for Spectrophotometry; Selection; Preparation; Certification; Use-SRM 930, NBS Spec. Publ. 260-51 (November 1975). $1.90* SN003-003-01481-8

Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials 730 and 799, from 4 to 3000 K, NBS Spec. Publ. 260-52 (September 1975). $1.05* SN003-003-01464-8

Durst, R. A., Standard Reference Materials: Standardization of pH Measurements, NBS Spec. Publ. 260-53 (December 1975, Revised). $1.05 SN003-003-01551-2

Burke, R. W., and Mavrodineanu, R., Standard Reference Materials: Certification and Use of Acidic Potassium Dichromate Solutions as an Ultraviolet Absorbance Standard, NBS Spec. Publ. 260-54 (August 1977). $3.00* SN003-003-01828-7

Ditmars, D. A., Cezairliyan, A., Ishihara, S., and Douglas, T. B., Standard Reference Materials: Enthalpy and Heat Capacity; Molybdenum SRM 781, from 273 to 2800 K, NBS Spec. Publ. 260-55 (September 1977). $2.20* SN003-003-01836-8

Powell, R. L., Sparks, L. L., and Hust, J. G., Standard Reference Materials: Standard Thermocouple Materials, Pt.67: SRM 1967, NBS Spec. Publ. 260-56 (February 1978). $2.20* SN003-003-018864

Cali, J. P. and Plebanski, T., Guide to United States Reference Materials, NBS Spec. Publ. 260-57 (February 1978). $2.20* PB 277173

Barnes, J. D., and Martin, G. M., Standard Reference Materials: Polyester Film for Oxygen Gas Transmission Measurements SRM 1470, NBS Spec. Publ. 260-58 (June 1979) $2.00* SN003-003-02077

Chang, T., and Kahn, A. H. Standard Reference Materials: Electron Paramagnetic Resonance Intensity Standard; SRM 2601, NBS Spec. Publ. 260-59 (August 1978) $2.30* SN003-003-01975-5

Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., and Moody, J. R., Standard Reference Materials: A Reference Method for the Determination of Sodium in Serum, NBS Spec. Publ. 260-60 (August 1978). $3.00* SN003-003 01978-0

Verdier, P. H., and Wagner. H. L., Standard Reference Materials: The Characterization of Linear Polyethylene (SRM 1482, 1483, 1484), NBS Spec. Publ. 260-61 (December 1978). $1.70* SN003-003-02006-1

Soulen, R. J., and Dove, R. B., Standard Reference Materials: Temperature Reference Standard for Use Below 0.5 K (SRM 768). NBS Spec. Publ. 260-62 (April 1979). $2.30* SN003-003-02047-8

Velapoldi, R. A., Paule, R. C., Schaffer, R. Mandel, J., Machlan, L. A., and Gramlich, J. W., Standard Reference Materials: A Reference Method for the Determination of Potassium in Serum. NBS Spec. Publ. 260-63 (May 1979). $3.75* SN003-003-02068

Velapoldi, R. A., and Mielenz, K. D., Standard Reference Materials: A Fluorescence Standard Reference Material Quinine Sulfate Dihydrate (SRM 936), NBS Spec. Publ. 260-64 (January 1980). $4.25* SN003-003-02148-2

Marinenko, R. B., Heinrich, K. F. J., and Ruegg, F. C., Standard Reference Materials: Micro-Homogeneity Studies of NBS Standard Reference Materials, NBS Research Materials, and Other Related Samples. NBS Spec. Publ. 260-65 (September 1979). $3.50* SN003-003-02114-1

Venable, W. H., Jr., and Eckerle, K. L., Standard Reference Materials: Didymium Glass Filters for Calibrating the Wavelength Scale of Spectrophotometers (SRM 2009, 2010, 2013). NBS Spec. Publ. 260-66 (October 1979). $3.50* SN003-003-02127-0

Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Murphy, T. J., and Gramlich, J. W., Standard Reference Materials: A Reference Method for the Determination of Chloride in Serum, NBS Spec. Publ. 260-67 (November 1979). $3.75* SN003-003-02136-9

Mavrodineanu, R. and Baldwin, J. R., Standard Reference Materials: Metal-On-Quartz Filters as a Standard Reference Material for Spectrophotometry-SRM 2031, NBS Spec. Publ. 260-68 (April 1980). $4.25* SN003-003-02167-9

Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Machlan, L. A., Garner, E. L., and Rains, T. C., Standard Reference Materials: A Reference Method for the Determination of Lithium in Serum, NBS Spec. Publ. 260-69 (July) 1980). $4.25* SN003-003-02214-4

Marinenko, R. B., Biancaniello, F., Boyer, P. A., Ruff, A. W., DeRobertis, L., Standard Reference Materials: Preparation and Characterization of an Iron-Chromium-Nickel Alloy for Microanalysis, NBS Spec. Publ. 260-70 (May 1981). $2.50* SN003-003-02328-1

Seward, R. W., and Mavrodineanu, R., Standard Reference Materials: Summary of the Clinical Laboratory Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-71 (November 1981). $6.50* SN003-003-02381-7

Reeder, D.J., Coxon, B., Enagonio, D., Christensen, R. G., Schaffer, R., Howell, B. F., Paule, R. C., Mandel, J., Standard Reference Materials: SRM 900, Antiepilepsy Drug Level Assay Standard, NBS Spec. Publ. 260-72 (June 1981). $4.25* SN003-003-02329-9

Interrante, C. G., and Hicho, G. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Fifteen Percent Austenite (SRM 486), NBS Spec. Publ. 260-73 (January 1982). $2.75* SN003-003-02386-8

Marinenko, R. B., Standard Reference Materials: Preparation and Characterization of K-411 and K-414 Mineral Glasses for Microanalysis: SRM 470. NBS Spec. Publ. 260-74 (April 1982). $3.50 SN003-003-023-95-7

Weidner, V. R., and Hsia, J. J., Standard Reference Materials: Preparation and Calibration of First Surface Aluminum Mirror Specular Reflectance Standards (SRM 2003a), NBS Spec. Publ. 260-75 (May 1982). $3.75 SN003-003-023-99-0

Hicho, G. E. and Eaton, E. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Five Percent Austenite (SRM 485a), NBS Spec. Publ. 260-76 (August 1982). $3.50 SN003-003-024-33-3

Furukawa, G. T., Riddle, J. L., Bigge, W. G., and Pfieffer, E. R., Standard Reference Materials: Application of Some Metal SRM's as Thermometric Fixed Points, NBS Spec. Publ. 260-77 (August 1982). $6.00 SN003-003-024-34-1

Hicho, G. E. and Eaton, E. E., Standard Reference Materials: Standard Reference Material Containing Nominally Thirty Percent Austenite (SRM 487), NBS Spec. Publ. 260-78 (September 1982). $3.75* SN003-003-024-35-0

Richmond, J. C., Hsia, J. J. Weidner, V. R., and Wilmering, D. B., Standard Reference Materials: Second Surface Mirror Standards of Specular Spectral Reflectance (SRM's 2023, 2024, 2025), NBS Spec. Publ. 260-79 (October 1982). $4.50* SN003-003-024-47-3

Schaffer, R., Mandel, J., Sun, T., Cohen, A., and Hertz, H. S., Standard Reference Materials: Evaluation by an ID/MS Method of the AACC Reference Method for Serum Glucose, NBS Spec. Publ. 260-80 (October 1982). $4.75* SN003-003-024-43-1

Burke, R. W., and Mavrodineanu, R. (NBS retired), Standard Reference Materials: Accuracy in Analytical Spectrophotometry, NBS Spec. Publ. 260-81 (April 1983). $6.00* SN003-003-024-8

Weidner, V. R., Standard Reference Materials: White Opal Glass Diffuse Spectral Reflectance Standards for the Visible Spectrum (SRM's 2015 and 2016). NBS Spec. Publ. 260-82 (April 1983). $3.75* SN003-003-024-89-9**

Bowers, G. N., Jr., Alvarez, R., Cali, J. P. (NBS retired), Eberhardt, K. R., Reeder, D. J., Schaffer, R., Uriano, G. A., Standard Reference Materials: The Measurement of the Catalytic (Activity) Concentration of Seven Enzymes in NBS Human Serum SRM 909, NBS Spec. Publ. 260-83 (June 1983). $4.50* SN003-003-024-99-6

Gills, T. E., Seward, R. W., Collins, R. J., and Webster, W. C., Standard Reference Materials: Sampling, Materials Handling, Processing, and Packaging of NBS Sulfur in Coal Standard Reference Materials, 2682, 2683, 2684, and 2685, NBS Spec. Publ. 260-84 (August 1983). $4.50* SN003-003-025-20-8

Swyt, D. A., Standard Reference Materials: A Look at Techniques for the Dimensional Calibration of Standard Microscopic Particles, NBS Spec. Publ. 260-85 (September 1983). $5.50* SN003-003-025-21-6

Hicho, G. E. and Eaton, E. E., Standard Reference Materials: A Standard Reference Material Containing Two and One-Half Percent Austenite, SRM 488, NBS Spec. Publ. 260-86 (December 1983). $1.75* SN003-003-025-41-1

Mangum, B. W., Standard Reference Materials: SRM 1969: Rubidium Triple-Point - A Temperature Reference Standard Near 39.30 °C, NBS Spec. Publ. 260-87 (December 1983). $2.25* SN003-003-025-44-5

Gladney, E. S., Burns, C. E., Perrin, D. R., Roelandts, I., and Gills, T. E., Standard Reference Materials: 1982 Compilation of Elemental Concentration Data for NBS Biological, Geological, and Environmental Standard Reference Materials. Spec. Publ. 260-88 (March 1984). $7.00* SN003-003-02565-8

Hust, J. G., Standard Reference Materials: A Fine-Grained, Isotropic Graphite for Use as NBS Thermophysical Property RM's from 5 to 2500 K, NBS Spec. Publ. 260-89 (September 1984). $4.50* SN003-003-02608-5

Hust, J. G., and Lankford, A. B., Standard Reference Materials: Update of Thermal Conductivity and Electrical Resistivity of Electrolytic Iron, Tungsten, and Stainless Steel, NBS Spec. Publ. 260-90 (September 1984). $3.00* SN003-003-02609-3

Goodrich, L. F., Vecchia, D. F., Pittman, E. S., Ekin, J. W., and Clark, A. F., Standard Reference Materials: Critical Current Measurements on an NbTi Superconducting Wire Standard Reference Material, NBS Spec. Publ. 260-91 (September 1984). $2.75* SN003-003-02614-0

Carpenter, B. S., Standard Reference Materials: Calibrated Glass Standards for Fission Track Use (Supplement to NBS Spec. 260-49). NBS Spec. Publ.. 260-92 (September 1984). $1.50* SN003-003-02610-7

Ehrstein, J., Preparation and Certification of Standard Reference Materials for Calibration of Spreading Resistance Probes, NBS Spec. Publ. 260-93 (January 1985). $2.25* SN003-003-02633-6

Gills, T. E., Koch, W. F. Stolz, J. W., Kelly, W. R., Paulsen, P. J., Colbert, J. C. Kirklin, D. R., Pei, P.T.S., Weeks, S., Lindstrom, R. M. Fleming, R. F., Greenberg, R. R., and Paule, R. C., Methods and Procedures Used at the National Bureau of Standards to Certify Sulfur in Coal SRM's for Sulfur Content, Calorific Value, Ash Content, NBS Spec. Publ. 260-94 (December 1984). $2.25* SN003-003-02629-8

Mulholland, G. W., Hartman, A. W., Hembree, G. G., Marx, E., and Lettieri, T. R., Standard Reference Materials: Development of a 1 m Diameter Particle Size Standard, SRM 1690, NBS Spec. Publ. 260-95 (May 1985). SN003-003-02665-4

Carpenter, B. S., Gramlich, J. W., Greenberg, R. R., and Machlan, L. A., Standard Reference Materials: Uranium-235 Isotopic Abundance Standard Reference Materials for Gamma Spectrometry Measurements, NBS Spec. Publ. 260-96 (In Preparation).

Mavrodineanu, R. and Gills, T. E., Standard Reference Materials: Summary of the Coal, Ore, Mineral, Rock, and Refractory Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-97 (September 1985).

Hust, J. G., Standard Reference Materials: Glass Fiberboard SRM for Thermal Resistance, NBS Spec. Publ. 260-98 (August 1985). SN003-003-02674-3

Callanan, J. E., Sullivan, S. A., and Vecchia, D. F., Standard Reference Materials: Feasibility Study for the Development of Standards Using Differential Scanning Calorimetry, NBS Spec. Publ. 260-99 (August 1985). SN003-003-02675-1

Taylor, J. K., Standard Reference Materials: Handbook for SRM Users, NBS Spec. Publ. 260-100 (In Press).

Mangum, B. W., Standard Reference Materials: SRM 1970, Succinonitrile Triple-Point Standard: A Temperature Reference Standard Near 58.08 °C, NBS Spec. Publ. 260-101 (In Preparation).

Mavrodineanu, R., Standard Reference Materials: Holmium Oxide Solution Wavelength Standard from 240- 650 nm, SRM 2034, NBS Spec. Publ. 260-102 (In Press).

Hust, J. G., Standard Reference Materials: Glass Fiberblanket SRM for Thermal Resistance, NBS Spec. Publ. 260-103 (September 1985). SN003-003-02687-5

Mavrodineanu, R. and Alvarez, R., Standard Reference Materials: Summary of the Biological and Botanical Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-104 (In Press).

* Send order with remittance to Superintendent of Documents, US Government Printing Office Washington, DC 20402. Remittance from foreign countries should include an additional one-fourth of the purchase price for postage.

** May be ordered from: National Technical Information Services (NTIS). Springfield Virginia 22161.

TABLE OF CONTENTS

Abstract

This publication is a summary of the biological and botanical Standard Reference Materials and Research Materials issued by the National Bureau of Standards. The material, composition, certification, use, and remarks concerning each of the ten materials described are presented in tabular form. Copies of the Certificates of Analysis for these materials are contained in the appendix for more detailed information.

Key Words: Biological materials; botanical materials; chemical compositions; Research Material; Standard Reference Material.

Introduction

Since its inauguration in 1901, the National Bureau of Standards (NBS) has issued nearly 2000 different Standard Reference Materials (SRM's). Many of these have been renewed several times, many have been replaced or discontinued as technology changed. Today, over 900 SRM's are available, together with a large number of scientific publications related to the fundamental and applied characteristics of these materials. Each material is certified for chemical composition, chemical properties, or its physical or mechanical characteristics. Each SRM is provided with a Certificate or Certificate of Analysis that contains the essential data concerning its properties or characteristics. The SRM's currently available cover a wide range of chemical, physical, and mechanical properties, and a corresponding wide range of measurement interests in practically all aspects of fundamental and applied science. These SRM's constitute a unique and invaluable means of transferring to the user accurate data obtained at NBS, and provide essential tools that can be used to improve accuracy in practically all areas where measurements are performed.

In addition to SRM's, the National Bureau of Standards issues a variety of Research Materials (RM's) having various properties described in individual "Reports of Investigation." They are intended primarily to further the scientific or technical research on that particular material. Other materials, called Special Reference Materials (GM's), are also available from NBS. These are materials produced and certified by other Government agencies, standard organizations, or other nonprofit organizations, that are considered useful to the public and for which no alternate method of national distribution exists.

The various categories of materials available from NBS are given in Table 1. This table lists these materials according to their chemical composition, physical properties, or engineering characteristics. A more detailed alphabetic enumeration of these materials is given in Appendix I. Table 1 and Appendix I were taken from NBS Special Publication 260, NBS Standard Reference Materials Catalog, 1984-85 Edition[1]. This publication lists every material available from the NBS Office of Standard Reference Materials.

Further information on the reference materials available from NBS may be obtained from the Office of Standard Reference Materials, National Bureau of Standards, Gaithersburg, MD 20899. Information on other NBS services may be obtained from the Technical Information and Publications Division, National Bureau of Standards, Gaithersburg, MD 20899.

In addition to these types of materials, NBS provides many additional services. These include: Measurement Assurance Programs, Calibration and Related Measurement Services, Proficiency Sample Programs, a National Voluntary Laboratory Accreditation Program, Standards Information Services, Standard Reference Data, and Technical Information and Publications.

---

[1]For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402, under Stock No. 003-003-02558-5 (Price $5.50, add 25 percent for foreign orders.)

Table 1. Categories of Standard Reference Materials available from the National Bureau of Standards.

CERTIFIED CHEMICAL COMPOSITION STANDARDS

Steels (chip form)

- Plain carbon
- Low alloy
- High alloy
- Stainless
- Tool

Steels (granular form)

Steels (solid form)

- Ingot iron and low alloy
- Special ingot irons and low alloy
- Stainless
- Specialty
- High-temperature alloys
- Tool

Steelmaking Alloys

Cast Irons (chip form)

Cast Steels, White Cast Irons, Ductile Irons, and Blast Furnace Irons (solid form)

Nonferrous Alloys (chip form)

- Aluminum "Benchmarks"
- Cobalt
- Copper
- Copper "Benchmarks"
- Lead
- Magnesium
- Nickel
- Nickel Superalloy, Trace Elements
- Nickel oxide
- Selenium
- Tin
- Titanium
- Zinc
- Zirconium

Nonferrous Alloys (solid form)

- Aluminum "Benchmarks"
- Copper
- Copper "Benchmarks"
- Lead
- Nickel
- Titanium
- Zinc
- Zirconium

Gases in Metals

High-Purity Metals

Electron Probe Microanalytical Standards

Primary, Working, and Secondary Standard Chemicals

Microchemical Standards

Clinical Laboratory Standards

Biological Standards

Environmental Standards

- Analyzed gases
- Analyzed liquids and solids
- Permeation tubes

Industrial Hygiene Standards

Forensic Standards

Hydrocarbon Blends

Metallo-Organic Compounds

Fertilizers

Ores

Minerals, Refractories, Glasses, and Carbides

Cement

Trace Element Standards

Nuclear Materials

- Special nuclear materials
  - Plutonium assay
  - Plutonium isotopic
  - Uranium assay
  - Uranium isotopic
- Neutron density standards
- Fission track glass standards

Isotopic Reference Standards

Table 1. continued.

CERTIFIED PHYSICAL PROPERTY STANDARDS

Ion Activity Standards

- pH standards
- pD standards
- Ion selective electrodes

Mechanical and Metrology Standards

- Magnification
- Coating thickness
- Glass
- Elasticity
- Density
- Polymer
- Rheology

Heat Standards

- Superconductive thermometric fixed point devices
- Freezing Points
  - Defining fixed points
  - Determined reference points
- Melting points
- Calorimetric
  - Combustion
  - Solution
  - Heat source
  - Enthalpy and heat capacity
- Vapor pressure
- Thermal expansion
- Thermocouple materials
- Thermal resistance

Magnetic Standards

- Magnetic susceptibility
- Magnetic moment
- Paramagnetic resonance

Optical Standards

- Spectrophotometric
- Thermal emittance
- Refractive index

Radioactivity Standards

- Alpha-particle standards
- Beta-particle and gamma-ray gas standards
- Alpha-particle, beta-particle, gamma-ray, and electron-capture solution standards
- Contemporary standard for carbon-14 dating laboratories
- Environmental standards
- Low energy photon sources
- Gamma-ray "point-source" standards
- Radium gamma-ray solution standards
- Radium solution standards for random analysis
- Radioactivity standard reference materials currently not in stock

Metallurgical

Mössbauer

X-ray Diffraction

Gas Transmission

Permittivity

Reference Fuels

Resistivity

ENGINEERING TYPE STANDARDS

Standard Rubber and Rubber-Compounding Materials

Reference Magnetic Tapes

Centerline Drawings, OCR-B

Sizing Standards

- Glass spheres for particle size
- Turbidimetric and fineness (cement)

RESEARCH MATERIALS

X-ray and Photographic Standards

Surface Flammability Standards

Smoke Density Chamber Standards

Water Vapor Permeance

Tape Adhesion Testing Standards

Color Standards

SPECIAL REFERENCE MATERIALS

The first NBS biological and botanical Standard Reference Material designed specifically for the analytical laboratory use was issued in January 1971. This standard, SRM 1571, Orchard Leaves, was developed in response to requests from numerous analytical chemists.

Since then, the number of biological and botanical SRM's has grown to 10, and during the next decade this growth is expected to continue at much the same rate.

This publication is an attempt to describe in general terms the composition, certification, and use of these biological and botanical SRM's.

Table 2 contains the essential information concerning the material, composition, the certification parameters, and use. Under "Remarks," additional data such as storage conditions and stability is provided.[1] All the data and information contained in this table were extracted from the Certificates of Analysis issued for the SRM's included in the table. An examination of this table gives the reader a general view of these SRM's. For more detailed information, the individual certificates reproduced in Appendix II should be consulted as well as the references cited in each certificate.

A summary of the quantitative distribution of the 47 chemical elements determined in the 9 Standard Reference Materials and 1 Research Material from Table 2 is presented in Table 3. The values with an asterisk are expressed in wt. %, the others are expressed in µg/g, and the data in parantheses are non-certified values.

The certificates in Appendix II are arranged in numerical order. The SRM's listed in the table include all of the biological and botanical standards that were issued or were in preparation by the end of 1984. These SRM's are the result of the concerted efforts of a number of scientists from the NBS National Measurement Laboratory as well as those from cooperating institutions. Each certificate lists the individuals who contributed to the certification of the SRM.

In addition to the SRM's and their certificates, NBS issues a series of Special Publications (SP), called the "260 Series," that relate directly to Standard Reference Materials as stated in the Preface. The list of available publications in the "260 Series" is given in the beginning of this publication.[2]

Other NBS publications, not in the "260 Series," and a number of NBS staff authored papers have been published that deal with specific SRM's or measurement techniques. Some of these are: SP 148, The Role of Standard Reference Materials in Measurement Systems; SP 378, Accuracy in Spectrophotometry and Luminescence Measurements (255 pp., 1973); and SP 466, Standardization in Spectrophotometry and Luminescence Measurments (150 pp., 1977), contains papers of interest to analytical chemists. Another publication that should be of particulate interest to the users of the SRM's described in Table 2 is SP 492, "Procedures Used at the National Bureau of Standards to Determine Selected Trace Elements in Biological and Botanical Materials."

This work consists of a collection of analytical procedures used in the Center for Analytical Chemistry, National Measurement Laboratory of the National Bureau of Standards for the determination of trace levels of Ag, Al, As, Be, Bi, Ca, Cd, Cr, Cu, Fe, Hg, K, Mg, Mn, Mo, Na, Ni, Pb, Pt, Sb, Se, Te, Tl, V, and Zn in biological and botanical materials. These procedures were critically selected or adapted, and often specially developed, by the scientific staff members of the Center for Analytical Chemistry to provide measurements with the best obtainable accuracy. They were considered to be most appropriate for the analysis and certification of various Standard Reference Materials issued by NBS such as SRM 1577a, Bovine Liver, SRM 1567, Wheat Flour, etc.

The description of these procedures is given with sufficient detail to permit the analyst to use them as a protocol for routine analyses in the laboratory.

They are assembled according to the analytical disciplines involved in the measuring process: sample preparation, neutron activiation analysis, spark source mass spectrometric isotope dilution, atomic absorption and flame emission spectrometry, molecular absorption spectrometry, fluorescence spectrometry, and polarography.

Special Publication 492 contains also a detailed description of the methods and instrumentation used to produce high-purity reagents, including the analytical techniques used to test their purity.

The important subject of blanks, a determining factor common to all analytical measurements, is examined including a detailed description of the environmental conditions necessary to control this essential factor, and to insure a maximum protection against contamination.

The lyophilization method for preconcentration and drying of various analytical samples is also discussed in this work.

The analytical procedures assembled in these publications are believed to provide the best measuring capabilities available at this time. They are, however, continuously being revised, improved, or replaced by more accurate ones.

Hence, it is recommended that the analyst who desires to be informed of these advancements maintain a contact with the scientific staff of the Center for Analytical Chemistry at NBS.

Another work, NBS Special Publication 422, "Accuracy in Trace Analysis: Sampling, Sample Handling, Analysis," Vol. I and II, P. D. LaFleur, Editor, issued in 1976 as a proceeding of the 7th Materials Research Symposium held at NBS in 1974, contains valuable information on the subject, to the analytical chemist.

---

[1] NOTE: The use of proprietary designations in Table 2 is for information only, and should not be construed as an endorsement of the product by either the Department of Commerce or the National Bureau of Standards.

[2] For complete bibliographic reference and ordering information, see "Other NBS Publications in This Series," pp. iv.

Table 2. Summary of the Biological and Botanical Standards

SRM	Material	Composition
RM50 Albacore Tuna	Tuna fish muscle tissue from albacore tuna caught in the San Diego, California area. It was frozen, ground, mixed, lyophilized, ground again, placed in polyethylene bags, and canned under nitrogen. To improve homogeneity the material was reground, re-blended and recanned.	Minor elements: K 1.22%, Na 0.11%. Trace elements (ppm): Hg 0.95; Se 3.6; Zn 13.6; As 3.3; Pb 0.46; Mn 1.3. Organic constituents (ppm): heptadiene 0.6; toluene 0.7; limonene 0.4; 2-nonanone 0.7; 2-undecanone 0.1; 2,6-di-t-butyl-p-cresol 1.0; hexa-decane, trace; heptadecane, trace; pristane, 0.03.
1549 Non-Fat Milk Powder	This material was obtained from a commercial source as a portion of a single lot. The moisture content of the material was reduced to a low level by the spray process.	Certified constituents: (wt. %) Ca 1.30; Cl 1.09; Mg 0.120; P 1.06; K 1.69; Na 0.497; S .351; (μg/g) Cd .0005; Cr .0026; Cu .7; Fe 1.78; I 3.38; Pb 0.019; Mn .26; Hg .0003; Se .11; Zn 46.1. Non-certified values given for 11 elements; also for lactose and ascorbic acid.
1566 Oyster Tissue	The oysters were obtained by the FDA Bureau of Shellfish Sanitation from a commercial source, frozen in sealed plastic bags. The material was ground, freeze-dried and powdered at the U.S. Army Natick Research & Development Command, Mass., blended and bottled at NBS and freeze-dried again.	Certified minor constituents: Ca 0.15%; Mg 0.128%; K 0.969%; Na 0.51%. Trace constituents (μg/g): As 13.4; Cd 3.5; Cr 0.69; Cu 63.0; Fe 195; Pb 0.48; Mn 17.5; Hg 0.057; Ni 1.03; Rb 4.45; Se 2.1; Ag 0.89; Sr 10.36; U 0.116; V 2.3; Zn 852. Non-certified information: Cl (1.0%); S (0.76%); P (0.81%); (μg/g) Br (55); Co (0.4); F (5.2); I (2.8); Mo (≤0.2); Tl (≤0.005); Th (0.1). Homo-geneity was determined by neutron activation and atomic absorption spectrometry on Na, Cl, V, Mn, Mg, K, Cu, Zn, Cd. Calcium exhibits some inhomogeneity.

Certification	Use	Remarks
RM 50 is not an SRM, hence, none of the data presented are certified. For further information see the Report of Investigation describing this Research Material.	RM 50 is intended to be used in the measurement of inorganic and organic chemical species in marine tissue at trace concentrations. It should be useful to scientists interested in evaluating analytical methods and interlaboratory comparisons.	RM 50, sealed in metal cans, should have an indefinite storage life under normal room conditions. The open sample can be kept for 6 months to 2 years in a polyethylene bag at 0 °C.
The analytical methods used in the certification were: atomic absorption spectrometry; atomic emission spectrometry; ion chromatography, isotope dilution mass spectrometry, neutron activation, and photon activation. All measurements are based on a minimum 500 mg of the dried material.	SRM 1549 is intended for use in calibrating instrumentation and evaluating the reliability of analytical methods for the determination of constituents in milk, milk powders, and other biological matrices.	The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation. The bottle should be kept tightly closed and stored in a desiccator in the dark. This certification is invalid after 3 years from date of shipping.
The analytical methods used in the certification were: atomic absorption, flame and spark emission spectrometry, inductively coupled plasma, thermal ionization and isotope dilution, isotope dilution mass spectrometry, spark source mass spectrometry; neutron activation, and polarography. All measurements are based on a minimum 250 mg of the dried material.	SRM 1566 is intended for calibrating instrumentation and validating methodology for the chemical analysis of marine animal tissue for minor and trace elements.	The material should be kept in its original bottle, stored at 10-30 °C in a dark desiccator. Under these conditions it should be stable for 5 years from date of shipping.

Table 2. Biological and Botanical Standards.

SRM	Material	Composition
1567 Wheat Flour	Wheat flour was milled from a blend of Hard Red Spring and Hard Red Winter wheat grown in South Dakota and subjected to 2.5 megarads of Co-60 radiation for microbiological control at Neutron Products, Inc., Dickerson, Md. The material was passed through a sieve with openings of 425 µm and blended.	Certified minor constituents: K 0.136%; Ca 0.019%. Trace constituents (µg/g): Fe 18.3; Zn 10.6; Mn 8.5; Na 8.0; Cu 2.0; Se 1.1; Cd 0.032; Hg 0.001. Non-certified information (µg/g): Br (9); Rb (1); Mo (0.4); Ni (0.18); As (0.006); Te (≤0.002). Homogeneity was determined by instrument neutron activation measurements on Mn, K, Zn, Na, and Br.
1568 Rice Flour	Rice flour was prepared from 100% long grain from Arkansas and subjected to 2.5 megarads of Co-60 radiation for microbiological control at Neutron Products, Inc., Dickerson, Md. The material was passed through a sieve with openings of 425 µm and blended.	Certified minor constituents: K 0.112%; Ca 0.014%. Trace constituents (µg/g): Mn 20.1; Zn 19.4; Fe 8.7; Na 6.0; Cu 2.2; As 0.41; Se 0.4; Cd 0.029; Co 0.02; Hg 0.0060. Non-certified information (µg/g): Rb (7); Mo (1.6); Br (1); Ni (0.16); Te (≤0.002) Homogeneity was determined by instrumental neutron activation measurements on Mn, K, Zn, Na, and Br.
1569 Brewers Yeast	Yeast was obtained from the Nutrition Institute, U.S. Dept. of Agriculture, Beltsville, Md. It was sieved (0.15 mm) and blended.	Cr, 2.12 ± 0.05 µg/g determined on the sample without drying and calculated on a dry weight basis. Homogeneity was determined by a neutron activation technique.
1572 Citrus Leaves	The plant material was collected from the Lake Alfred area of central Florida and prepared at Michigan State University. It was air-dried, ground to pass a 425 µm sieve, dried at 85 °C and mixed. It was sterilized with Co-60 radiation at the U.S. Army Research & Development Command, Natick, Mass.	Certified major and minor constituents (wt. % on dry weight basis): Ca 3.15; Mg 0.58; P 0.13; K 1.82; S 0.407. Trace constituents (µg/g): Al 92; As 3.1; Ba 21; Cd 0.03; Cr 0.8; Cu 16.5; I 1.84; Fe 90; Pb 13.3; Mn 23; Hg 0.08; Mo 0.17; Ni 0.6; Rb 4.84; Na 160; Sr 100, Zn 29. Non-certified values (µg/g): Sb (0.04); Br (8.2); Ce (0.28); Cs (0.098); Cl (414); Co (0.02) Eu (0.01); La (0.19); Sm (0.052); Sc (0.01); Se (0.025); Te (0.02) Tl (≤0.01); Sn (0.24); U (≤0.15); N (2.86 wt. %).

Certification	Use	Remarks
The analytical methods used in the certification were: atomic absorption, flame emission, and isotope dilution spark source mass spectrometry, neutron activation, and polarography. All measurements are based on a minimum 400 mg sample and are reported on a "dry-weight" basis. Selenium and mercury are determined on the material without drying.	SRM 1567 is intended for calibrating instrumentation and evaluating the reliability of analytical methods for the determination of minor and trace elements in wheat flour and similar food products.	The material should be kept in its original bottle, stored at 10-30 °C in a dark desiccator. Under these conditions it should be stable for 5 years from date of shipping.
Same as SRM 1567.	Same as SRM 1567.	Same as SRM 1567.
Chromium was determined by neutron activation and by isotope dilution mass spectromtry. The analyses were performed on the sample without drying and the results expressed on a dry weight basis established on separate samples heated at 85 °C for 3 hrs. The samples should not be dissolved in open vessels.	SRM 1569 is intended for use in calibrating instrumentation and evaluating the accuracy of analytical methods for the determination of chromium in Brewers Yeast and other biological materials containing a volatile chromium component which presents an especially difficult problem.	The material should be kept in its original bottle tightly capped, and stored at 10-23 °C in a dark desiccator.
The analytical methods used in the certification were: atomic absorption and flame emission spectrometry, atomic emission spectrometry using inductively coupled plasma, ion chromatography, isotope dilution thermal source mass spectrometry, isotope dilution spark source mass spectrometry, Kjeldahl method for nitrogen, neutron activation, photon activation, polarography, spectrophotometry.	SRM 1572 is intended for use in calibrating instrumentation and evaluating the analytical methods used for the determination of major, minor, and trace elements in botanical materials, agricultural food products, and similar matrices.	The material should be kept tightly closed in its original bottle, stored in a dark desiccator at a temperature between 10-30 °C. A minimum sample of 500 mg of the dried material should be used for analyses.

Table 2. Biological and Botanical Standards.

SRM	Material	Composition
1573 Tomato Leaves	The leaves were obtained from a field of direct seeded tomatoes at the Horticultural Research Center of the Michigan State University. The air-dried leaves were ground, dried at 85 °C, mixed, packaged in polyethylene-lined drums, and sterilized with Co-60 radiation at the U.S. Army Research & Development Command, Natick, Mass.	Certified major and minor constituents (wt. %): K 4.46; Ca 3.00; P 0.34. Trace constituents (μg/g): Fe 690; Mn 238; Zn 62; Sr 44.9; Rb 16.5; Cu 11; Pb 6.3; Cr 4.5; As 0.27; Th 0.17; U 0.061. Non-certified information, major and minor constituents (wt. %): N (5.0); Mg (0.7); Al (0.12). Trace constituents (μg/g): B (30); Br (26); Cd (3); Ce (1.6); La (0.9); Co (0.6); Sc (0.13); Hg (0.1); Tl (0.05); Eu (0.04) Homogeneity was evaluated by determining P, Fe, Mn, Zn, Rb, Cu, Cr, As, and U.
1575 Pine Needles	Collected from the Manistee State Park, Mich., Air-dried, ground, and dried again at 85 °C. It was mixed, passed through a 0.25 mm sieve, packaged in polyethylene drums, and sterilized with Co-60 radiation at the U.S. Army Research & Development Command, Natick, Mass.	Certified minor elements: Ca 0.41%; K 0.37%; P 0.12%. Trace constituents (μg/g): Mn 675; Al 545; Fe 200; Rb 11.7; Pb 10.8; Sr 4.8; Cu 3.0; Cr 2.6; As 0.21; Hg 0.15; Th 0.037; U 0.020. Non-certified information: N (1.2%); and Br (9); Ni (3.5) Ce (0.4); Cd (<0.5); Sb (0.2); La (0.2); Co (0.1); Tl (0.05); Sc (0.03); Eu (0.006) μg/g. Homogeneity was established by determining P, Al, Fe, Mn, Rb, Cu, Cr, As, Hg, U, K, Ca, Sr, Pb, and Th.
1577a Bovine Liver	The liver was obtained in the Portland, Oregon area. It was ground, mixed, and lyophilized in polyethylene trays by Oregon Freeze Dry Food Inc., Albany, Oregon. The material was then powdered and packaged in moisture-proof bags.	Certified elements (wt. %): Cl 0.28; P 1.11; K 0.996; Na 0.243; S 0.78. Trace elements (μg/g): As 0.047; Cd 0.44; Ca 120; Co 0.21; Cu 158; Fe 194; Pb 0.135; Mg 600; Mn 9.9; Hg 0.004; Mo 3.5; Rb 12.5; Se 0.71; Ag 0.04; Sr 0.138; U 0.00071; V 0.099; Zn 123. Non-certified information: N (10.7 wt. %); and (μg/g); Al (2); Sb (0.003), Br (9); Tl (0.003). Homogeneity was tested by analyzing randomly selected samples and found to be satisfactory.

Certification	Use	Remarks
The analytical methods used in the certification were: atomic absorption and optical emission spectroscopy, isotope dilution and spark source mass spectrometry, Kjeldahl, neutron activation, nuclear track technique, spectrophotometry, polarography. A minimum of 500 mg of the dried material should be used for the analysis.	SRM 1573 is intended for calibrating instrumentation and validation of analytical methods for the determination of major, minor, and trace elements in botanical materials and other agricultural products.	The material should be kept in the original bottle stored at 10-30 °C in a dark desiccator. Under these conditions SRM 1573 should be stable for 5 years after shipping date.
The analytical methods used in the certification were: atomic absorption spectroscopy, isotope dilution mass spectrometry, isotope dilution spark source mass spectrometry, Kjeldahl method, neutron activiation, nuclear track technique, optical emission spectroscopy, spectrophotometry, polarography. A minimum of 500 mg of the dried material should be used for the analysis.	Same as SRM 1573.	Same as SRM 1573.
The analytical methods used in the certification were: atomic absorption and flame emission spectrometry, ion chromatogrphy, inductively coupled plasma emission spectrometry, isotope dilution thermal and spark source mass spectrometry, Kjeldahl method for nitrogen, neutron activation, polarography, and spectrophotometry. A minimum of 250 mg of the dried material should be used for the analysis.	SRM 1577a is intended for calibration of instrumentation and validation of methods in the chemical analysis of animal tissue for major, minor, and trace elements.	Same as SRM 1573.

Table 3. Summary of the Analytical Data Obtained for the Biological and Botanical Standard Reference Materials and Research Material.

SRM	1549	1566	1567	1568	1569	1572	1573	1575	1577a	RM 50†
	Non-Fat Milk Powder	Oyster Tissue	Wheat Flour	Rice Flour	Brewers Yeast	Citrus Leaves	Tomato Leaves	Pine Needles	Bovine Liver	Albacore Tuna
Element										
Ag	(<0.0003)	0.89	----	----	----	----	----	----	0.04	----
Al	(2)	----	----	----	----	92	(0.12)*	545	(2)	----
As	(0.0019)	13.4	(0.006)	0.41	----	3.1	0.27	0.21	0.047	(3.3)
B	----	----	----	----	----	----	(30)	----	----	----
Ba	----	----	----	----	----	21	----	----	----	----
Br	(12)	(55)	(9)	(1)	----	(8.2)	(26)	(9)	(9)	----
Ca	1.30*	0.15*	0.019*	0.014*	----	3.15*	3.00*	0.41*	120	----
Cd	0.0005	3.5	0.032	0.029	----	0.03	(3)	(<0.5)	0.44	----
Ce	----	----	----	----	----	(0.28)	(1.6)	(0.4)	----	----
Cl	1.09*	(1.0)*	----	----	----	(414)	----	----	0.28*	----
Co	(0.0041)	(0.4)	----	0.02	----	(0.02)	(0.6)	(0.1)	0.21	----
Cr	0.0026	0.69	----	----	2.12	0.8	4.5	2.6	----	----
Cs	----	----	----	----	----	(0.098)	----	----	----	----
Cu	0.7	63.0	2.0	2.2	----	16.5	11	3.0	158	----
Eu	----	----	----	----	----	(0.01)	(0.04)	(0.006)	----	----
F	(0.20)	(5.2)	----	----	----	----	----	----	----	----
Fe	1.78	195	18.3	8.7	----	90	690	200	194	----
Hg	0.0003	0.057	0.001	0.0060	----	0.08	(0.1)	0.15	0.004	(0.95)
I	3.38	(2.8)	----	----	----	1.84	----	----	----	----
K	1.69*	0.969*	0.136*	0.112*	----	1.82*	4.46*	0.37*	0.996	(1.22)*
La	----	----	----	----	----	(0.19)	(0.9)	(0.2)	----	----
Mg	0.120*	0.128*	----	----	----	0.58*	(0.7)*	----	600	----
Mn	0.26	17.5	8.5	20.1	----	23	238	675	9.9	(1.3)
Mo	(0.34)	(≤0.2)	(0.4)	(1.6)	----	0.17	----	----	3.5	----
N	----	----	----	----	----	(2.86)*	(5.0)*	(1.2)*	(10.7)*	----

SRM	1549	1566	1567	1568	1569	1572	1573	1575	1577a	RM 50†
	Non-Fat Milk Powder	Oyster Tissue	Wheat Flour	Rice Flour	Brewers Yeast	Citrus Leaves	Tomato Leaves	Pine Needles	Bovine Liver	Albacore Tuna
Element										
Na	0.497*	0.51*	8.0	6.0	----	160	----	----	0.243*	(0.11)*
Ni	----	1.03	(0.18)	(0.16)	----	0.6	----	(3.5)	----	----
P	1.06	(0.81)*	----	----	----	0.13*	0.34*	0.12*	1.11*	----
Pb	0.019	0.48	0.020	0.045	----	13.3	6.3	10.8	0.135	(0.46)
Rb	(11)	4.45	(1)	(7)	----	4.84	16.5	11.7	12.5	----
S	0.351*	(0.76)*	----	----	----	0.407*	----	----	0.78*	----
Sb	(0.00027)	----	----	----	----	(0.04)	----	(0.2)	(0.003)	----
Sc	----	----	----	----	----	(0.01)	(0.13)	(0.03)	----	----
Se	0.11	2.1	1.1	0.4	----	(0.025)	----	----	0.71	(3.6)
Si	(<50)	----	----	----	----	----	----	----	----	----
Sm	----	----	----	----	----	(0.052)	----	----	----	----
Sn	(<0.5)	----	----	----	----	(0.24)	----	----	----	----
Sr	----	10.36	----	----	----	100	44.9	4.8	0.138	----
Te	----	----	(≤0.002)	(≤0.002)	----	(0.02)	----	----	----	----
Th	----	(0.1)	----	----	----	----	0.17	0.037	----	----
Tl	----	(≤0.005)	----	----	----	(≤0.01)	(0.05)	(0.05)	(0.003)	----
U	----	0.116	----	----	----	(≤0.15)	0.061	0.020	0.00071	----
V	----	2.3	----	----	----	----	----	----	0.099	----
Zn	46.1	852	10.6	19.4	----	29	62	----	123	(13.6)

REMARKS

*The certified data with an asterisk are expressed in wt. %; those in parentheses are non-certified values. The remaining certified data are expressed in µg/g.

†This is not an SRM, hence the values given for this Research Material are not certified. Several organic compounds are also identified.

Appendix I.

Alphabetical Index by Standard Reference Material Name

Name	SRM
Acetanilide	141c
Acid Open-Hearth Steel, 0.2% Carbon	19G
Acid Potassium Phthalate	84j
AISI 1045 Steel	20g
AISI 4340 Steel	361
AISI 4340 Steel	1261a
AISI 94B17 Steel (Modified)	362
AISI 94B17 Steel (Modified)	1262a
Albacore Tuna	RM 50
Alkali Lead Silicate Glass	712
Alpha Quartz	1878
Alumina (Reduction Grade)	699
Alumina Silicate Glass	714
Aluminosilicate Glass	715
Aluminum Alloy	85B
Aluminum Alloy 6011 (Modified)	858
Aluminum Alloy 6011 (Modified)	1258
Aluminum Alloy 7075	859
Aluminum Alloy 7075	1259
Aluminum Block, Eddy Current Conductivity	1860
Aluminum Block, Eddy Current Conductivity	1861
Aluminum Block, Eddy Current Conductivity	1862
Aluminum Block, Eddy Current Conductivity	1863
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	1118
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	C1118
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	1119
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	C1119
Aluminum Casting Alloy 356	855
Aluminum Casting Alloy 380	856
Aluminum Cube Ultra Purity	RM 1C
Aluminum 2-Ethylhexanoate	1075a
Aluminum, Freezing Point Standard	44f
Aluminum, Magnetic Gram Susceptibility	763
Aluminum Oxide, Melting Point	742
Aluminum Rod Ultra Purity	RM 1R
Aluminum-26 Radioactivity Standard	4229
Americium-241 Alpha-Particle Standard	4904F
Americium-241 Gamma-ray Standard	4213
Ammonium Dihydrogen Phosphate	194
Angiotensin I (Human)	998
Anisic Acid	142
Anticonvulsant Drug Level Assay Standard	1599
Antiepilepsy Drug Level Assay Standard	900
Antimony-125-Tellurium-125m, Europium-154, Europium-155 Mixed-Radionuclide Point-Source Standard	4275B
Antimony-125-Tellurium-125m, Europium-154, Europium-155 Mixed-Radionuclide Solution Standard	4276B
A.O.H., 0.4C Spectrographic Steel Standard	413
Argillaceous Limestone	1C
Arsenic Trioxide Reductometric Standard	83d
Assay-Isotopic Standard for Potassium	985
Assay-Isotopic Standard for Rhenium	989
Assay-Isotopic Standard for Silicon	990
Assay-Isotopic Standard for Strontium	987
2% Austenite in Ferrite	488
5% Austenite in Ferrite	485a
15% Austenite in Ferrite	486
30% Austenite in Ferrite	487

Name	SRM
Austenitic Stainless Steel, Thermal Conductivity and Electrical Resistivity	1460
Austenitic Stainless Steel, Thermal Conductivity and Electrical Resistivity	1461
Austenitic Stainless Steel, Thermal Conductivity and Electrical Resistivity	1462
Barium Crown Glass	713
Barium Cyclohexanebutyrate	1051b
Barrium-133 Radioactivity Point-Source Standard	4241B
Barium-133 Radioactivity Standard	4251B
Basalt Rock	688
Base Oil	1083
Basic Electric Spectrographic Steel Standard	404a
Basic Open-Hearth Steel, 0.1% Carbon	15g
Basic Open-Hearth Steel, 0.1% Carbon	335
Basic Open-Hearth Steel, 0.1% Carbon	1228
Basic Open-Hearth Steel, 0.2% Carbon	11h
Basic Open-Hearth Steel, 0.4% Carbon	12H
Basic Open-Hearth Steel, 0.5% Carbon	152A
Basic Open-Hearth Steel, 0.8% Carbon	14f
Basic Open-Hearth Steel, 1% Carbon (Disk)	1227
Basic Open-Hearth Steel, 1.1% Carbon	16f
Basic Open-Hearth Steel, 1.1% Carbon	337
0.4C Basic Oxygen Furnace Steel	178
Bauxite (Arkansas)	69b
Bauxite (Dominican)	697
Bauxite (Jamaican)	698
Bauxite (Surinam)	696
Benzene in Nitrogen	1805
Benzene in Nitrogen	1806
Benzene Permeation Device	1911
Benzoic Acid	140b
Benzoic Acid	350a
Benzoic Acid Calorimetric Standard	39i
Benzothiazyl Disulfide Rubber Compound	373f
Beryllium-Copper Standard	1122
Beryllium-Copper Standard	C1122
Beryllium-Copper Standard	C1123

Name	SRM
Beryllium on Filter Media	2675
Bessemer Steel (Simulated) 0.1% Carbon	8j
Bilirubin	916
Bis(1-phenyl-1, 3-butanediono) copper (II)	1080a
Bis(1-phenyl-1, 3-butanediono) oxovanadium (IV)	1052b
Black Porcelain Enamel for Directional Hemispherical Reflectance	2021
Black Porcelain Enamel for Directional Hemispherical Reflectance	2022
Blast Furnace Iron Standard (Chill Cast White)	1143a
Blast Furnace Iron Standard (Chill Cast White)	1144a
B.O.H., 0.4C Spectrographic Steel Standard	417a
Boric Acid	951
Boron-Doped Silicon Slices for Resistivity Measurements	1521
Borosilicate Glass	93a
Borosilicate Glass	623
Borosilicate Glass	717
Borosilicate Glass	1825
Borosilicate Glass, Thermal Expansion	731
Bovine Liver	1577a
Bovine Serum Albumin	926
Bovine Serum Albumin (7% Solution)	927
Branched Polyethylene	1476
Brewers Yeast	1569
Bright Copper Microhardness Standard	1894
Bright Nickel Microhardness Standard	1895
Bromobenzoic Acid	2142
Burnt Refractory	76a
Burnt Refractory	77a
Burnt Refractory	78a
Cadmium Cyclohexanebutyrate	1053a
Cadmium, Vapor Pressure	746
Calcium Carbonate	915
Calcium 2-Ethylhexanoate	1074a
Calcium in Low-Alloy (Silicon) Steel	1254
Calcium Molybdate	71
Calibrated Glass Beads	1004
Calibrated Glass Beads	1017a
Calibrated Glass Beads	1018a
Calibrated Glass Spheres	1003a
Carbon Dioxide in Air	1670
Carbon Dioxide in Air	1671
Carbon Dioxide in Air	1672
Carbon Dioxide in Nitrogen	1674b
Carbon Dioxide in Nitrogen	1675b
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2619a
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2620a

Name	SRM
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2621a
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2622a
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2623a
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2624a
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2625a
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2626a
Carbon Dioxide in Nitrogen (Mobile Source Emission Gas Standard)	2632
Carbon Dioxide in Nitrogen (Mobile Source Emission Gas Standard)	2633
Carbon Monoxide in Air (Ambient Air Quality Gas Standard)	2612a
Carbon Monoxide in Air (Ambient Air Quality Gas Standard)	2613a
Carbon Monoxide in Air (Ambient Air Quality Gas Standard)	2614a
Carbon Monoxide in Nitrogen	1677c
Carbon Monoxide in Nitrogen	1678c
Carbon Monoxide in Nitrogen	1679c
Carbon Monoxide in Nitrogen	1680b
Carbon Monoxide in Nitrogen	1681b
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2635
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2636
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2637
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2638
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2639
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2640
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2641
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2642
Carbon-14 Radioactivity Standard	4245
Carbon-14 Radioactivity Standard	4246
Carbon Steel	1224
Carbon Steel, 0 6%	13g
Cast Iron	4k
Cast Iron	5L
Cast Iron	6g
Cast Iron	7G
Cast Iron Car Wheel	122h
Cast Steel 3	C1173
Cast Steel Standard	1138a
Cast Steel Standard	1139a

Name	SRM
Catalyst Package for Lubricant Oxidation	1817
Centerline Drawings for Optical Character Recognition, B Characters	1901
Centroid Color Chart	2106
Centroid Color Kit	2107
Cesium-137, Barium-137m Point-Source Radioactivity Standard	4200B
Cesium-137, Barium-137m Point-Source Radioactivity Standard	4207
Cesium-137 Burn-Up Standard	4233B
Cesium-134 Radioactivity Standard	4250B
Channel Black Rubber Compound	375g
Chlorine-36 Beta-ray Standard	4943
Chlorine-36 Radioactivity Standard	4422L
Chlorobenzoic Acid	2144
Chrome Refractory	103a
Chromium-Molybdenum-Aluminum Steel	106B
Chromium-Molybdenum Steel	36b
Chromium-Molybdenum Steel	133B
Chromium-Nickel-Molybdenum Steel	139b
Chromium-Nickel-Molybdenum Steel	1222
17Chromium-9 Nickel-0.2 Selenium Steel	339
Chromium-Nickel Spectrographic Steel Standard	408a
15Chromium-7 Nickel Steel	344
16Chromium-4 Nickel Steel	345
Chromium-51 Radioactivity Standard	4400L-F
Chromium Steel	163
Chromium-Tungsten Steel	155
Chromium-Vanadium Spectrographic Steel Standard	407a
Cholesterol	911a
Chrysotile Asbestos Fibers	1876
Citrus Leaves	1572
Clinical Laboratory Thermometer	934
Cobalt Cyclohexanebutyrate	1055b
Cobalt-Molybdenum-Tungsten Steel	153A
Cobalt-57 Radioactivity Standard	4408L-C
Cobalt-60 Radioactivity Standard	4915D
Commerical Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	1115
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	C1115
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	1116

Name	SRM
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	C1116
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	1117
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	C1117
Common Lead Isotopic Standard	981
Copper Concentrate	332
Copper Heat Capacity Test Specimen	RM5
Copper-Nickel-Chromium Cast Iron	115A
Copper Ore, Mill Heads	330
Copper Ore, Mill Tails	331
Copper-Thermal Expansion	736a
Copper, Secondary Freezing Point Standard	45d
Cortisol (Hydrocortisone)	921
Creatinine	914
Cr-Mo Low Alloy Steel	1270
Cr-Mo Steel (ASTM A-213)	291
Cr-Mo (SAE 4140) Spectrographic Steel Standard	414
Cr-Mo (SAE 4150) Spectrographic Steel Standard	427
Cr-Mo (SAE X4130) Spectrographic Steel Standard	418a
Cr-Ni-Mo Steel (AISI 8620)	293
18Cr-10Ni Steel (AISI 304L)	101f
Cr-V Steel (Modified)	363
Cr-V Steel (Modified)	1263a
Cr-V Steel (SAE 6150)	30f
Crystalline Potassium Dichromate	935
Crystalline Potassium Iodide, Heterochromatic Stray Radiant Energy Standard	2032
Crystalline (Ruby) Electron Paramagnetic Resonance Absorption Intensity Standard	2601
Cupro-Nickel (CDA 706)	1275
Cupro-Nickel (CDA 715)	1276
Cupro-Nickel, 10% (CDA 706) Doped	875

Name	SRM
Cupro-Nickel, 10% (CDA 706) High Purity	874
Cystine	143c
Dextrose	41b
D-Glucose	917
Dibutyltin Bis(2-ethylhexanoate)	1057b
Didymium Glass Filter for Checking the Wavelength Scale of Spectrophotometers	2009
Didymium Glass Fitler for Checking the Wavelength Scale of Spectrophotometers	2010
Disodium Hydrogen Phosphate	186IIc
Disodium Hydrogen Phosphate	2186II
D-Mannitol	920
Dolomitic Limestone	88a
Doped Platinum	681L1
Doped Platinum	681L2
Ductile Cast Iron	341
Electrical Residual Resistivity Ratio Standard	769
Electrolytic Iron	365
Electrolytic Iron	1265a
Electrolytic Iron, Thermal Conductivity and Electrical Resistivity	1463
Electrolytic Iron, Thermal Conductivity and Electrical Resistivity	1464
Electronic and Magnetic Alloy Standard	1159
Electronic and Magnetic Alloy Standard	1160
Enriched Boric Acid	952
Equal-Atom Lead Isotopic Standard	982
Estuarine Sediment	1646
Europium-152 Point-Source Standard	4218E
Europium-152 Radioactivity Standard	4370B
Extra Dense Lead Glass	709
Fe-Cr-Ni Alloy Microprobe Standard	479a
Fe-3Si Alloy Microprobe Standard	483
Feldspar	70a
Feldspar	99a
Ferrochromium (Low Carbon)	196
Ferrochromium Silicon	689
Ferroniobium	340
Ferrophosphorus	90
Ferrosilicon	58a
Ferrosilicon	59a
Ferrosilicon (75% Si)	195
First Surface Aluminum Mirror for Specular Reflectance	2003a
First Surface Mirror, Gold on Glass	2008a

Name	SRM
Fission Track Glass Standard	961
Fission Track Glass Standard	962a
Fission Track Glass Standard	963a
Fission Track Glass Standard	964
Flint Clay	97a
Fluorobenzoic Acid	2143
Fluorspar	79a
Free-Cutting Brass	1103
Free-Cutting Brass	C1104
Freeze-Dried Urine	2670
Freeze-Dried Urine Certified for Fluoride	2671a
Freeze-Dried Urine Certified for Mercury	2672a
Fused-Silica Thermal Expansion	739
Gadolinium-148 Alpha-Particle Standard	4907
Gallium Melting-Point Standard	1968
Gallium-67 Radioactivity Standard	4416L-D
Gas Furnace Black Rubber Compound	382a
Gasometric Set (1095-1099)	1089
Gasometric Standard for Unalloyed Zirconium	357
Gasometric Standard for Unalloyed Zirconium	358
Generator Columns for Polynuclear Aromatic Hydrocarbons	1644
Gilding Metal	1112
Gilding Metal	C1112
Gilding Metal	1113
Gilding Metal	C1113
Gilding Metal	1114
Gilding Metal	C1114
Glasses for Microchemical Analysis	1871
Glasses for Microchemical Analysis	1872
Glasses for Microchemical Analysis	1873
Glasses for Microchemical Analysis	1874
Glasses for Microchemical Analysis	1875
Glass Fibers for Microanalysis	RM 31
Glass Filter for Transmittance Measurement	2030
Glass Filters for Spectrophotometry	930D
Glass Fluorescence Source	477
Glass Sand	81a
Glass Sand	165a
Glass Spheres	1019a
Gold Coating on Glass Sealing Alloy	1398a
Gold Coating on Nickel	1379
Gold Coating on Nickel	1380
Gold Coating on Nickel	1399b
Gold-Copper Wires for Microprobe Analysis	482
Gold-195 Radioactivity Standard	4421L

Name	SRM
Gold-198 Radioactivity Standard	4405L-B
Gold-Silver Wires for Microprobe Analysis	481
Gold, Vapor Pressure	745
Gray Cast Iron	334
Halocarbons (in methanol) for Water Analysis	1639
High-Alloy Steel (A-743)	C1288
High-Alloy Steel (AISI 310 Mod.)	C1287
High-Alloy Steel, (AISI 414 Mod.)	C1289
High-Alloy White Cast	892
High-Alloy White Cast Iron	890
High-Alloy White Cast Iron	891
High-Carbon Ferrochromium	64c
High-Carbon Ferromanganese	68c
High-Carbon Steel (Modified)	364
High-Carbon Steel (Modified)	1264a
High-Grade Fluorspar	180
High-Nickel Steel	126c
High-Nickel Steel	1158
High-Purity Gold	685
High-Purity Platinum	680L1A
High-Purity Platinum	680L2A
High-Purity Platinum Thermoelement	1967
High-Purity Zinc	682
High-Silicon Steel	179
High-Silicon Steel	1134
High-Silicon Steel	1135
High-Silicon Steel (Calcium Bearing)	125b
High-Sulfur Steel	105
High-Sulfur Steel	129c
High-Sulfur Steel	1136
High Temperature Alloy A286	348
High Temperature Alloy M308	1197
High Temperature Alloy L605 and S816	S1199
High-Temperature Alloy	1206-2
High-Temperature Alloy	1207-1
High-Temperature Alloy	1207-2
High-Temperature Alloy	1208-1
High-Temperature Alloy	1208-2
Homogeneous River Sediment for Radioactivity Measurements	RM 45B
Human Liver, Environmental Radioactivity	4352
Human Lung, Environmental Radioactivity	4351
Human Serum	909

Name	SRM
Hydrogen in Unalloyed Titanium	352b
Hydrogen in Unalloyed Titanium	1086
Hydrogen in Unalloyed Titanium	1087
Hydrogen in Unalloyed Titanium	1088
Hydrogen-3 Radioactivity Standard	4361
Hydrogen-3 Radioactivity Standard	4926C
Hydrogen-3 Toluene Radioactivity Standard	4947
4-Hydroxy-3 methoxy-DL-mandelic Acid (VMA)	925
ICTA High Temperature Set Differential Thermal Analysis	GM 760
ICTA Low Temperature Set Differential Thermal Analysis	GM 757
ICTA Mod Temperature Set Differential Thermal Analysis	GM 759
ICTA Mid Temperature Set Differential Thermal Analysis	GM 758
ICTA Polystyrene Differential Thermal Analysis	GM 754
ICTA Thermogravimetry Set	GM 761
Incoloy, 901 and Hastelloy X	S1198
Inconels, Alloy 600 (Chips)	864
Inconels, Alloy 600 (Solid)	1244
Inconels, Alloy 625 (Chips)	865
Inconels, Alloy 625 (Solid)	1245
Incoloy, Alloy 800 (Chips)	866
Incoloy, Alloy 800 (Solid)	1246
Incoloy, Alloy 825 (Chips)	867
Incoloy, Alloy 825 (Solid)	1247
Indium-111 Radioactivity Standard	4417L-C
Ingot Iron Spectrographic Steel Standard	420a
Intermediate Purity Selenium	726
Intermediate-Purity Zinc	728
Iodine-123 Radioactivity Standard	4414L-C
Iodine-125 Radiactivity Standard	4407L-H
Iodine-129 Radioactivity Standard	4949B
Iodine-131 Radioactivity Standard	4401L-I
Iron Foil Mössbauer Standard	1541
Iron-55 Low-Energy Photon Standard	4260C
Iron Metal (Clinical Standard)	937
Iron Ore (Labrador)	692
Iron Ore (Nimba)	693

Name	SRM
Iron Ore (Sibley)	27f
Iron Ore Concentrate (Canada)	690
Iron-59 Radioactivity Standard	4411L-B
Isobutylene-Isoprene (Butyl) Rubber	1495
Isobutylene-Isoprene (Butyl) Rubber	388L
Isotopic Standard for Bromine	977
Isotopic Standard for Chlorine	975
Isotopic Standard for Chromium	979
Isotopic Standard for Copper	976
Isotopic Standard for Magnesium	980
Isotopic Standard for Silver	978
Krypton-85 Gaseous Radioactivity Standard	4308C
Krypton-85 Radioactivity Standard	4235
Krypton-85 Radioactivity Standard	4935C
Lead-Barium Glass	89
Lead-Base Bearing Metal	53e
Lead-Base Bearing Metal	1132
Lead Cyclohexanebutyrate	1059c
Lead in Reference Fuel	1636a
Lead in Reference Fuel	1637a
Lead in Reference Fuel	1638a
Lead Nitrate	928
Lead on Filter Media	2674
Lead-203 Radioactivity Standard	4420L
Lead, Secondary Freezing Point Standard	49e
Lead-Silica Glass	1827
Lead-Silica Glass (Viscosity)	711
Lead-Silica Glass for dc Volume Resistivity	624
Lead-Silica Glass for Dielectric Constant	774
Lead 206 Spike Assay and Isotopic Solution Standard	991
Leaded-Tin Bronze Alloy	1035
Light-Sensitive Paper	700d
Light-Sensitive Paper	701d
Light-Sensitive Plastic Chip	703
Linear Polyethylene	1475
Linear Polyethylene	1482
Linear Polyethylene	1483
Linear Polyethylene	1484
Linerboard, Standard for Tape Adhesion Testing	1810
Liquid Absorbance Standard for Ultraviolet and Visible Spectrophotometry	931c
Lithium Carbonate	924
Lithium Ore	181
Lithium Ore	182
Lithium Ore	183
Low-Alloy Steel, (AISI 4130)	1225
Low Alloy Steel	1226
Low Alloy Steel (A242 Mod.)	C1285
Low-Alloy Steel, AISI 4130	72g
Low Alloy Steel (AISI 1526, Modified)	1269
Low-Alloy Steel (Hy 80)	1286

Name	SRM
Low-Alloy Steel Set (661-665)	S668
Low-Carbon Silicon Steel	131c
Low-Carbon Silicon Steel	1036
Low-Carbon Stainless Steel (AISI 316L)	166c
Magnesium-base Alloy	171
Magnesium Cyclohexanebutyrate	1061c
Magnesium Gluconate Dihydrate	929
Magnetic Coating on Magnetic Substrate (Nickel on Steel)	1365a
Magnetic Coating on Magnetic Substrate (Nickel on Steel)	1366a
Magnetic Coating on Non-Magnetic Substrate (Nickel and Chromium on brass	1367a
Magnetic Tape, High Density	6250
Manganese Fluoride, Magnetic Gram Susceptibility	766
Manganese Ore	25d
Manganese-54 Point-Source Radioactivity Standard	4997E
Manganese-54 Radioactivity Standard	4257
Manganese Steel	100B
Manganous Cyclohexanebutyrate	1062b
Maraging Steel	1156
Metal on Quartz Filters for Spectrophotometry	2031
Metals on Filter Media	2676b
Methane in Air	1658a
Methane in Air	1659a
Methane in Air	1660a
Medium Manganese Spectrographic Steel Standard	405a
Mercaptobenzothiazole	383a
Mercury, Freezing Point	743
Mercury-203 Radioactivity Standard	4418L
Mercury in Water, μg/mL	1641b
Mercury in Water, ng/mL	1642b
Microcopy Resolution Test Chart	1010a
Microprobe Standard – Cartridge Brass	478
Mineral Glasses for Microanalysis	470
Molybdenum Concentrate	333
Molybdenum, Heat Capacity	781
Molybdenum-99 Radioactivity Standard	4412L-H
Molybdenum-Tungsten-Chromium-Vanadium Steel	134A
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	1106
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	C1106
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	1107
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	C1107

Name	SRM
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	1108
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	C1108
Neutral Glass	716
Neutron Density Monitor Wire	953
Nickel-Chromium Cast Iron	82b
Nickel-Chromium-Molybdenum Cast Iron	107C
Nickel-Chromium Steel	32E
Nickel-Copper Alloy	882
Nickel Cyclohexanebutyrate	1065b
Nickel Oxide, No. 1	671
Nickel Oxide, No. 2	672
Nickel Oxide, No. 3	673
Nickel-63 Radioactivity Standard	4226
Nickel Silver (CDA 762)	879
Nickel Siver (CDA 770)	880
Nickel Spectrographic Steel Standard	409b
Nickel Sphere, Magnetic Moment	772
Nickel Steel	33d
Ni-Cr-Mo-V Steel	1173
Nicotinic Acid	148
Niobium-94 Gamma-ray Standard	4201B
Nitric Oxide in Nitrogen	1683b
Nitric Oxide in Nitrogen	1684b
Nitric Oxide in Nitrogen	1685b
Nitric Oxide in Nitrogen	1686b
Nitric Oxide in Nitrogen	1687b
Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2627
Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2628
Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2629
Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2630
Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2631
Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2653
Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2654
Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2655
Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2656
Nitrogen Dioxide Permeation Device	1629a
4-Nitrophenol	938

Name	SRM
Nodular Cast Iron	342a
Nominal One Micrometer Polystyrene Spheres	1690
Non-Fat Powdered Milk	1549
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1359
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1360
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1361b
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1362a
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1363a
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1364a
NPL GM Alpha Alumina	8005
NPL GM Alpha Alumina	8006
NPL GM Alpha Alumina	8007
NPL GM Alpha Alumina	8008
NPL GM Graphitized Carbon Black	8001
NPL GM Graphitized Carbon Black	8002
NPL GM Melting Point Set	8000
NPL GM Non-porous Silica	8003
NPL GM Non-porous Silica	8004
N-tertiary-Butyl-2-benzothiazolesulfen-amide Rubber Compound	384d
Obsidian Rock	278
Octaphenylcyclotetrasiloxane	1066a
Oil Furnace Black Rubber Compound	378b
Opal Glass Powder	91
Optical Emission and X-ray Spectroscopic Analysis	1102
Optical Microscope Linewidth Measurement Standard	474
Optical Microscope Linewidth Measurement Standard	475
Optical Microscope Linewidth Measurement Standard	476

Name	SRM
Organics in Shale Oil	1580
Oxalic Acid	4990C
Oxygen in Ferrous Materials Ingot Iron	1090
Oxygen in Ferrous Materials (Stainless Steel AISI 431)	1091
Oxygen in Ferrous Materials Vacuum Melted Steel	1092
Oxygen in Maraging Steel	1094
Oxygen in Nitrogen (Gas Standard)	2657
Oxygen in Nitrogen (Gas Standard)	2658
Oxygen in Nitrogen (Gas Standard)	2659
Oxygen in Titanium-Base Materials	355
Oxygen in Valve Steel	1093
Oyster Tissue	1566
Palladium, Magnetic Gram Susceptibility	765
Penetrant Test Block	1850
Peruvian Soil, Environmental Radioactivity	4355
Petroleum Crude Oil	1582
Phosphate Rock (Florida)	120b
Phosphor Bronze (CDA 521)	871
Phosphor Bronze (CDA 544)	872
Phosphorized Copper, Cu VIII	C1251
Phosphorized Copper, Cu IX	C1252
Phosphorized Copper, Cu X	C1253
Phosphorus-32 Radioactivity Standard	44061 G
Photographic Step Tablet	1008
Pine Needles	1575
Plastic Clay	98a
Platinum, Magnetic Gram Susceptibility	764
Plutonium-238 Alpha-Particle Standard	4906B
Plutonium-240 Alpha-Particle Emission-Rate Solution Standard	4338
Plutonium-239 Alpha-Particle Solution Standard	4331
Plutonium-242 Alpha-Particle Solution Standard	4334B
Plutonium Isotopic Standard	946
Plutonium Isotopic Standard	947
Plutonium Isotopic Standard	948
Plutonium Metal	949f
Plutonium Metal (Standard Matrix Material)	945
Plutonium-244 Spike Assay and Isotopic Standard	996
Polychlorinated Biphenyls in Oil	1581
Polycrystalline Alumina Elasticity Standard	718
Polyester Plastic Film for Oxygen Gas Transmission	1470
Polyisobutylene Solution in Cetane	1490
Polystyrene	1478
Polystyrene	1479
Polystyrene (Broad Molecular Weight)	706
Polystyrene (Narrow Molecular Weight)	705
Polystyrene Spheres	1691
Portland Cement (Black)	1880

Name	SRM
Portland Cement (Blue)	635
Portland Cement (Clear)	639
Portland Cement (Gold)	634
Portland Cement (Green)	638
Portland Cement (Pink)	637
Portland Cement (Red)	633
Portland Cement (White)	1881
Portland Cement (Yellow)	636
Portland Cement Fineness Standard	114n
Potassium Chloride	2202
Potassium Chloride (Clinical Standard)	918
Potassium Chloride (Primary Chemical)	999
Potassium Chloride for Solution Calorimetry	1655
Potassium Dichromate	136d
Potassium Dihydrogen Phosphate	200
Potassium Dihydrogen Phosphate	186Ic
Potassium Dihydrogen Phosphate	2186I
Potassium Erucate	1076
Potassium Feldspar	607
Potassium Fluoride	2203
Potassium Hydrogen Phthalate	185e
Potassium Hydrogen Tartrate	188
Potassium Iodide with Attenuator	2033
Potassium Nitrate	193
Potassium Tetroxalate	189
Powdered Lead Based Paint	1579
Priority Pollutant Polynuclear Aromatic Hydrocarbons (in Acetonitrile)	1647
Propane in Air	1665b
Propane in Air	1666b
Propane in Air	1667b
Propane in Air	1668b
Propane in Air	1669b
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2643
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2644
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2645
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2646
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2647
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2648
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2649
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2650
Propane in Nitrogen and Oxygen (Mobile Source Emission Gas Standard)	2651
Propane in Nitrogen and Oxygen (Mobile Source Emission Gas Standard)	2652
Quartz Cuvette for Spectrophotometry	932
Quartz for Hydrofluoric Acid Solution Calorimetry	1654

Name	SRM
Quartz on Filter Media	2679a
Quinine Sulfate Dihydrate	936
Radiogenic Lead Isotopic Standard	983
Radium-226 Gamma-ray Standard	4956
Radium-226 Gamma-ray Standard	4957
Radium-226 Gamma-ray Standard	4958
Radium-226 Gamma-ray Standard	4959
Radium-226 Gamma-ray Standard	4960
Radium-226 Gamma-ray Standard	4961
Radium-226 Gamma-ray Standard	4962
Radium-226 Gamma-ray Standard	4963
Radium-226 Gamma-ray Standard	4964B
Radium Standard (Blank Solution)	4952B
Radon-226 for Radon Analysis	4953C
Red Brass	1109
Red Brass	C1109
Red Brass	1110
Red Brass	C1110
Red Brass	1111
Red Brass	C1111
Reduced Iron Oxide	691
Reference Fuel Isooctane	1816a
Reference Fuel n-Heptane	1815a
Reflection Step Tablet	2061
Refractive Index Glass	1820
Refractive Index Silicone Liquids	1823
Refractive Index, Soda-Lime Glass	1822
Relative Stress-Optical Coefficient Glass	708
Resulfurized-Rephosphorized Steel	C1221
Rice Flour	1568
River Sediment	1645
River Sediment, Environmental Radioactivity	4350B
Rocky Flats Soil Number 1, Environmental Radioactivity	4353
Rubidium Melting Point	1969
Rutile Ore	670
Scanning Electron Microscope Magnification Standard	484c
Scanning Electron Microscope Performance Standard	2069
Secondary Standard Flexible Disk Cartridge (Computer Amplitude Reference)	3210
Secondary Standard Magnetic Tape	3200
Secondary Standard Magnetic Tape Cassette	1600
Secondary Standard Magnetic Tape Cartridge (Computer Amplitude Reference)	3216
Second Surface Aluminum Mirror for Specular Reflectance	2023

Name	SRM
Second Surface Aluminum Mirror for Specular Reflectance	2024
Second Surface Aluminum Mirror with Wedge for Specular Reflectance	2025
Selenium-Bearing Steel	1170b
Selenium-75 Radioactivity Standard	4409L-D
Sheet Brass	37E
Silica Brick	198
Silica Brick	199
Silicon-Aluminum Alloy	87a
Silicon Bronze	158A
Silicon Density Standard	1840
Silicon Density Standard	1841
Silicon Metal	57a
Silicon Powder, Spacing Standard for X-ray Diffraction	640a
Silicon Power Device Level Resistivity Standard	1522
Silicon Resistivity Standard for Eddy Current Testers	1523
Silver 2-Ethylhexanoate	1077a
Silver-Gold Thermocouple Wire	733
Silver, Vapor Pressure	748
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1465
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1466
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1467
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1468
Smoke Density Chamber Standard (Flaming Exposure Condition)	1007a
Smoke Density Chamber Standard (Non-flaming Exposure Condition)	1006b
Soda-Lime Container Glass	621
Soda-Lime Flat Glass	620
Soda-Lime Float Glass	1830
Soda-Lime Glass	1826
Soda-Lime Glass Powder	92

Name	SRM
Soda-Lime Sheet Glass	1831
Soda-Lime Silica Glass	622
Soda-Lime Silica Glass	710
Soda-Lime Silica Glass for Liquidus Temperature	773
Sodium Bicarbonate	191a
Sodium Bicarbonate	2191
Sodium Carbonate	192a
Sodium Carbonate	2192
Sodium Chloride	2201
Sodium Chloride (Clinical Standard)	919
Sodium Cyclohexanebutyrate	1069b
Sodium Oxalate Reductometric Standard	40h
Sodium Pyruvate	910
Sodium Tetraborate Decahydrate (Borax)	187b
Solder	127b
Solder	1131
Special Nuclear Container DOT 6M, 15 gal.	9940
Special Nuclear Container, 55 gal.	9941
Special Nuclear Container Type A, 10 gal.	9942
Special Nuclear Container, Type A, 55 gal.	9943
Special Nuclear Material Package	9910
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	461
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	462
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	463
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	464
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	465
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	466
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	467
Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	468
Spectrograhic Ingot Iron and Low-Alloy Steel Standard	1166
Spectrographic Stainless Steel Standard	442
Spectrographic Stainless Steel Standard	443
Spectrographic Stainless Steel Standard	444
Spectrographic Stainless Steel Standard (Disc)	D849
Spectrographic Stainless Steel Standard (Disc)	D850
Spectrographic Stainless Steel Standard (Group II)	445

Name	SRM
Spectrographic Stainless Steel Standard (Group II)	446
Spectrographic Stainless Steel Standard (Group II)	447
Spectrographic Stainless Steel Standard (Group II)	448
Spectrographic Stainless Steel Standard (Group II)	449
Spectrographic Stainless Steel Standard (Group II)	450
Spectrographic Stainless Steel Standard (Rod)	849
Spectrographic Stainless Steel Standard (Rod)	850
Spectrographic Steel Standard (Disc)	D803a
Spectrographic Steel Standard (Disc)	D807a
Spectrographic Steel Standard (Rod)	803a
Spectrographic Steel Standard (Rod)	804a
Spectrographic Steel Standard (Rod)	805a
Spectrographic Steel Standard (Rod)	807a
Spectrographic Steel Standard (Rod)	808a
Spectrographic Steel Standard (Rod)	809a
Spectrographic Steel Standard (Rod)	817b
Spectrographic Steel Standard (Rod)	820a
Spectrographic Steel Standard (Rod)	821
Spectrographic Steel Standard (Rod)	827
Spectrographic Tool Steel Standard	436
Spectrographic Tool Steel Standard	437
Spectrographic Tool Steel Standard	438
Spectrographic Tool Steel Standard	439
Spectrographic Tool Steel Standard	440
Spectrographic Tool Steel Standard	441
Spectrographic Tool Steel Standard	837
Spectrographic Tool Steel Standard	840
Spectrographic Tool Steel Standard (Disc)	D837
Spectrographic Tool Steel Standard (Disc)	D840
Spectrographic Tool Steel Standard (Disc)	D841
Spectrographic Zinc-Base Die-Casting Alloy A	625
Spectrographic Zinc-Base Die-Casting Alloy B	626
Spectrographic Zinc-Base Die-Casting Alloy C	627
Spectrographic Zinc-Base Die-Casting Alloy D	628
Spectrographic Zinc-Base Die-Casting Alloy E	629
Spectrographic Zinc-Base Die-Casting Alloy F	630
Spectrographic Zinc Spelter Standard	631
Spectroscopic Titanium-Base Standard	641
Spectroscopic Titanium-Base Standard	642
Spectroscopic Titanium Base Standard	643

Name	SRM
Spectroscopic Titanium-Base Standard	644
Spectroscopic Titanium-Base Standard	645
Spectroscopic Titanium-Base Standard	646
Spheroidized Iron Carbide in Ferrite	493
Spreading Resistance Calibration (100) n-Type Silicon	2529
Spreading Resistance Calibration (100) p-Type Silicon	2528
Spreading Resistance Calibration (111) n-Type Silicon	2527
Spreading Resistance Calibration (111) p-Type Silicon	2526
Stabilized Wine	1590
Stainless Steel	121d
Stainless Steel	123c
Stainless Steel	160b
Stainless Steel (AISI 446)	367
Stainless Steel (AISI 446)	1267
Stainless Steel, 13% Chromium	73c
Stainless Steel, Cr-Ni	C1151
Stainless Steel, Cr-Ni	1151a
Stainless Steel, Cr-Ni	C1152
Stainless Steel, Cr-Ni	1152a
Stainless Steel, Cr-Ni	C1153
Stainless Steel, Cr-Ni	1153a
Stainless Steel, Cr-Ni	C1154
Stainless Steel, Cr-Ni	1154a
Stainless Steel, Cr-Ni-Mo	1155
Stainless Steel, Cr-Ni-Nb	1172
Stainless Steel, Cr-Ni-Ti	1171
Stainless Steel for Pitting or Crevice Corrosion	1890
Stainless Steel Thermal Expansion	738
Stearic Acid Rubber Compound	372h
Steel (AISI 1211)	368
Steel (Lead-Bearing)	1169b
Strontium Cyclohexanebutyrate	1070a
Strontium-85 Radioactivity Standard	4403L-B
Strontium-89 Radioactivity Standard	4945D
Styrene-butadiene Rubber (Type 1500)	386h
Succinonitrile Freezing Point	1970
Sucrose	17c
Sulfate and Nitrate on Filter Media	2673
Sulfur Dioxide in Nitrogen	1661a
Sulfur Dioxide in Nitrogen	1662a
Sulfur Dioxide in Nitrogen	1663a
Sulfur Dioxide in Nitrogen	1664a
Sulfur Dioxide in Nitrogen	1693
Sulfur Dioxide in Nitrogen	1694

Name	SRM
Sulfur Dioxide in Nitrogen	1696
Sulfur Dioxide Permeation Tube (2 cm tube)	1627
Sulfur Dioxide Permeation Tube (5 cm tube)	1626
Sulfur Dioxide Permeation Tube (10 cm tube)	1625
Sulfur in Coal	2682
Sulfur in Coal	2683
Sulfur in Coal	2684
Sulfur in Coal	2685
Sulfur in Residual Fuel Oil	1619
Sulfur in Residual Fuel Oil	1620a
Sulfur in Residual Fuel Oil	1621b
Sulfur in Residual Fuel Oil	1622b
Sulfur in Residual Fuel Oil	1623a
Sulfur in Residual Fuel Oil	1624a
Sulfur Rubber Compound	371g
Superconductive Thermometric Fixed Point Device	767a
Superconductive Thermometric Fixed Point Device	768
Surface Flammability Standard	1002c
Synthetic Sapphire	720
Technetium-99 Radioactivity Standard	4288
Technetium-99m Radioactivity Standard	4410H-I
Tetrachloroethylene in Nitrogen	1808
Thallium-201 Radioactivity Standard	4404L-F
Thermal Resistance, Fibrous Glass Batt	1451
Thermal Resistance, Fibrous Glass Board	1450b
Thorium-228, Thallium-208 Gamma-ray Point-Source Standard	4206C
Tin-Base Bearing Metal	54D
Tin, Freezing Point	741
Tin-113-Indium-113m Radioactivity Standard	4402L-C
Tin-121m Point-Source Gamma-ray Emission-Rate Standard	4264B
Tin, Secondary Freezing Point Standard	42g
Titanium Alloy	654a
Titanium-Base Alloy	173b
Titanium-Base Alloy	176

Name	SRM
Titanium-Base Alloy (Unalloyed)	650
Titanium-Base Alloy (Unalloyed)	651
Titanium-Base Alloy (Unalloyed)	652
Titanium Dioxide	154b
Toluene	211c
Tomato Leaves	1573
Tool Steel (AISI M2)	132b
Tool Steel (AISI M2)	1157
Tool Steel Abrasive Wear Standard	1857
Tracealloy (Nickel-Base High-Temperature Alloy)	897
Tracealloy (Nickel-Base High-Temperature Alloy)	898
Tracealloy (Nickel-Base High-Temperature Alloy)	899
Trace Elements in a Glass Matrix	610
Trace Elements in a Glass Matrix	611
Trace Elements in a Glass Matrix	612
Trace Elements in a Glass Matrix	613
Trace Elements in a Glass Matrix	614
Trace Elements in a Glass Matrix	615
Trace Elements in a Glass Matrix	616
Trace Elements in a Glass Matrix	617
Trace Elements in Coal (Bituminous)	1632a
Trace Elements in Coal (Sub-bituminous	1635
Trace Elements in Coal Fly Ash	1633a
Trace Elements in Fuel Oil	1634a
Trace Elements in Water	1643a
Trace Mercury in Coal	1630
2,2,4-Trimethylpentane	217c
Tripalmitin	1595
Tris, Basimetric	723a
Tris, for Solution Calorimetry	724a
Tris(hydroxymethyl)aminomethane	922
Tris(hydroxymethyl)aminomethane hydrochloride	923
Tris(1-phenyl-1, 3-butanediono) Chromium (III)	1078b
Tris(1-phenyl-1, 3-butanediono) Iron (III)	1079b
Triphenyl Phosphate	1071b
Tungsten Carbide	276a
Tungsten-Chromium-Vanadium Steel	50c
Tungsten Concentrate	277
Tungsten, Heat Capacity	782
Tungsten-20% Molybdenum Alloy Electron Microprobe Standard	480
Tungsten Thermal Expansion	737
Unalloyed Copper	1034
Unalloyed Copper, Cu "O"	393
Unalloyed Copper, Cu IV	457
Unalloyed Copper, Cu XI	454
Unalloyed Copper, Cu I (Chip)	394
Unalloyed Copper, Cu II (Chip)	395
Unalloyed Copper, Cu III (Chip)	396
Unalloyed Copper, Cu V (Chip)	398
Unalloyed Copper, Cu VI (Chip)	399
Unalloyed Copper, Cu VII (Chip)	400
Unalloyed Copper, Cu I (Rod)	494

Name	SRM
Unalloyed Copper, Cu II (Rod)	495
Unalloyed Copper, Cu III (Rod)	496
Unalloyed Copper, Cu V (Rod)	498
Unalloyed Copper, Cu VI (Rod)	499
Unalloyed Copper, Cu VII (Rod)	500
Unalloyed Titanium	354
Uranium Isotopic Standard (Nominally depleted to 0.02%)	U-0002
Uranium Isotopic Standard	U-005a
Uranium Isotopic Standard (Nominally 1% Enriched)	U-010
Uranium Isotopic Standard (Nominally 1.5% Enriched)	U-015
Uranium Isotopic Standard	U-020
Uranium Isotopic Standard	U-030a
Uranium Isotopic Standard (Nominally 5% Enriched)	U-050
Uranium Isotopic Standard (Nominally 10% Enriched)	U-100
Uranium Isotopic Standard (Nominally 15% Enriched)	U-150
Uranium Isotopic Standard (Nominally 20% Enriched)	U-200
Uranium Isotopic Standard (Nominally 35% Enriched)	U-350
Uranium Isotopic Standard (Nominally 50% Enriched)	U-500
Uranium Isotopic Standard (Nominally 75% Enriched)	U-750
Uranium Isotopic Standard (Nominally 80% Enriched)	U-800
Uranium Isotopic Standard (Nominally 85% Enriched)	U-850
Uranium Isotopic Standard (Nominally 90% Enriched)	U-900
Uranium Isotopic Standard (Nominally 93% Enriched)	U-930
Uranium Isotopic Standard (Nominally 97% Enriched)	U-970
Uranium Metal	960
Uranium Oxide	950b
Uranium Oxide	969
Uranium-233 Spike Assay and Isotopic Solution Standard	995
Uranium-235 Spike Assay and Isotopic Solution Standard	993
Urban Dust/Organics	1649
Urban Particulate Matter	1648
Urea	912a
Urea	2141
Urea	2152
Uric Acid	913
Vanadium and Nickel in Residual Fuel Oil	1618
Vanadium in Curde Oil	8505
Vanadium-49 Low-Energy Photon Standard	4266
Waspaloy	349
Wear-Metals in Lubricating Oil (100 ppm)	1084

Name	SRM
Wear-Metals in Lubricating Oil (300 ppm)	1085
Wheat Flour	1567
White Cast Iron	338
White Cast Iron (Disc)	1145
White Cast Iron (Disc)	1146
White Cast Iron (Disc)	1150
White Ceramic Tile for Directional Hemispherical Reflectance	2019b
White Ceramic Tile for Directional Hemispherical Reflectance	2020
White Iron	3d
White Opan Glass Diffuse Spectral Reflectance Standard for the Visible Spectrum	2015
Xenon-127 Gaseous Radioactivity Standard	4309G
Xenon-133 Gaseous Radioactivity Standard	4307I
Xenon-133 Gaseous Radioactivity Standard	4415L-I
Xenon-133, Xenon-137, Krypton-85 Mixed Gaseous Radioactivity Standard	4310B
X-ray Film Step Tablet	1001
X-ray Powder Diffraction Intensity Standard	674
X-ray Powder Diffraction (Mica) Low 2 Theta	675
Ytterbium-169 Radioactivity Standard	4419L-B
Zinc-Base Alloy (Die Casting)	94c
Zinc Concentrates	113a
Zinc Concentrates	329
Zinc Cyclohexanebutyrate	1073b
Zinc, Freezing Point	740
Zinc, Freezing Point Standard	43h
Zinc Metal	683
Zinc Oxide Rubber Compound	370e
Zircaloy-2	360a
Zircaloy-4 Metal	1237
Zircaloy-4 Metal	1238
Zircaloy-4 Metal	1239
Zirconium-Barium Chromate Formulation for Heat-Source Powder Calorimetry	1651
Zirconium-Barium Chromate Formulation for Heat-Source Powder Calorimetry	1652
Zirconium-Barium Chromate Formulation for Heat-Source Powder Calorimetry	1653
Zirconium Metal	1234
Zirconium Metal	1235
Zirconium Metal	1236

U.S. Department of Commerce
Juanita M. Kreps
Secretary
National Bureau of Standards
Ernest Ambler, Acting Director

National Bureau of Standards
Report of Investigation
Research Material 50
Albacore Tuna

P. D. LaFleur and W. P. Reed

A lyophilized (freeze-dried) marine biological tissue sample has been prepared in an attempt to satisfy many of the analytical requirements for a base line marine reference material.

Tuna fish muscle tissue was chosen for this purpose because of availability, use as a fish foodstuff, and its oceanographic interest.

The tuna is available as a Research Material (RM), in sets of two cans. Each can contains approximately 35 grams of lyophilized tuna tissue in a polyethylene bag, inside the hermetically-sealed, nitrogen filled can.

Material Application

This particular RM is intended to be used in the measurement of elements present at trace concentrations. In addition, some measurements of trace hydrocarbons have been made. The material represents a typical marine tissue that has been lyophilized. It should prove useful to those scientists who may wish to evaluate analytical methods, or to use a generally available "real" sample matrix in interlaboratory comparisons. It is important to note, however, that this is not a Standard Reference Material and none of the data presented here are certified. For the convenience of the analyst using this material, we have included a discussion for each component reported. It is apparent that there is significant heterogeneity for some elements, specifically the "bone-seeking" elements. The homogeneity of other elements appears to be acceptable. In most instances the heterogeneity of the material is observed between samples from the same can. There is some evidence that this heterogeneity is due primarily to fine particles of cartilage or bone present with the tissue in this RM.

Material Preparation

The tuna tissue used in this research material is from albacore tuna caught in the San Diego area in July of 1971. The tuna was cleaned, filleted, frozen and transported to a lyophilization facility. It was thawed, ground, and mixed using stainless steel equipment. For the final mixing, the entire lot was held in a single stainless steel container. The tuna was then lyophilized in aluminum trays lined with polyethylene. After lyophilization, the material was again ground in a stainless-steel mill, carefully transferred to new, individual polyethylene bags, and canned under nitrogen for storage.

Preliminary studies performed on these canned samples raised serious problems about the homogeneity of the material. In an effort to improve this condition the material was reground, reblended and recanned under the same conditions, as before. Except where stated, all measurements provided in this report were made on the reprocessed material.. During the second regrinding, fibrous material tended to "float" to the top due to the action of the mixer. This material was discarded.

(over)

Material Use

The lyophilized tuna tissue, sealed in metal cans, should have an indefinite storage life under normal room conditions. Thus far, no evidence exists to indicate deterioration of the material with time as long as the can remains sealed. Once opened, the possibility exists that the raw material will turn rancid. Researchers using only a portion of the material have successfully stored the remainder by placing the polyethylene bag in a glass jar with lid and storing at or below 0°C. A shelf life of 6 months is not unusual under these circumstances and a 2 year stability has been reported. The freeze-dried material represents only about 30% of the original weight of the tuna tissue. Consequently, for direct comparisons with fresh tissue the researcher may wish to adjust the sample weight for this difference. Please note, however, that all values in this "Report of Investigation" are based on the lyophilized weight of the tuna tissue.

The dissolution of this material is not difficult and researchers have reported using a variety of techniques. These techniques include nitric and perchloric acid digestion, nitric-sulfuric-perchloric (4:1:1) acid digestion, low temperature ashing, and oxygen combustion. Because of the apparent presence of small bits of cartilage and/or bone in the sample, the sample size for analysis should be 250 mg or greater to obtain reproducible results. For "bone-seeking" elements (e.g., Pb, U, Ca, Sr) the sample may be nonhomogeneous even with much larger sample sizes.

For the elements listed below, sufficient analytical work has been performed to permit some evaluation of the data. These brief evaluations are not certified values, but only judgments as to the amount of the element present. Problems encountered by analysts making these measurements are also described.

Mercury

The question of the mercury content of various foodstuffs has been studied by many investigators It is hoped that this material will provide a base line RM for environmental studies of mercury in food. A value of 0.95 ± 0.1 ppm encompasses the means of all of the reported values for mercury with one exception and extensive studies have indicated good homogeneity for this element.

Agreement among methods appears to be good. The data show very little bias among the three methods used. Sample sizes for these measurements have been 0.25 grams and larger.

The question of volatile mercury was explored by one investigator who reported the mercury content decreasing after opening the container. This decrease amounted to approximately 0.1 ppm over a period of 3 weeks after opening even though stored at −25 °C. This work has not been confirmed. Further work by several investigators has suggested that 80-90% of the mercury content is present as methylmercury.

Selenium

This element was determined by four laboratories using both neutron activation analysis and atomic absorption spectroscopy. The range of reported values ($\bar{x}$) is from 3.27 to 4.01 ppm. A most probable value is 3.6 ± 0.4. There has been very little indication of heterogeneity for this element.

Zinc

The analysis for zinc was performed by three laboratories using two analytical methods, neutron activation analysis and atomic absorption analysis. The range of mean values ($\bar{x}$) is 11.4 to 14.6 ppm. Our estimate of the probable value is 13.6 ± 1 ppm. No random behavior has been noted in the analysis of this element.

Arsenic

The analysis for arsenic was performed by 4 laboratories using both neutron activation and atomic absorption techniques. Data obtained within each laboratory appears to be consistent, although there appears to be some disagreement among laboratories analyzing for this element. The range of average values ($\bar{x}$) between laboratories is 2.74 to 4.6 ppm. The recommended value for the arsenic content is 3.3 ± 0.4 ppm.

Lead

There is a limited amount of data available for lead. The average value is 0.46 ppm. The homogeneity of the material for this element is questionable and the range of individual values is quite large. The use of this Research Material as a control, or for the development of methods for lead cannot be recommended. The isotopic composition, however, has been determined and is: ^{208}Pb 52.2%; ^{207}Pb 21.5%; ^{206}Pb 24.9%; and ^{204}Pb 1.38%.

Other Elements

Manganese was found to be distributed homogeneously and the measured value in the original lot of tuna (not the reblended material) was 1.3 ppm. Sodium also was found to be homogeneous in the original lot of material and the reported value was 0.11%.

Potassium was found to be distributed homogeneously and the measured value on the original lot of tuna (not the reblended material) was 1.22%. The following elements have been found to be heterogeneously distributed in the tuna tissue: uranium, thorium, calcium, and strontium.

Organic Materials

The tuna research material was subjected to headspace sampling and analysis. The concentrations given in the following table are to be considered only as order-of-magnitude estimates of the hydrocarbon compounds present, as they were calculated relative to the amount of naphthalene added to the sample and most of the compounds identified were not aromatic hydrocarbons.

The value for the largest constituent, 2,6-di-t-butyl-p-cresol, is an especially poor estimate. However, this compound is interesting because it is a common antioxidant used in food packaging. It probably has its origins in one of the handling steps between tuna collection and packaging in plastic. An indication of petroleum pollution in the tuna sample comes from the aliphatic hydrocarbons present.

The identification of the monoterpene, limonene, is reasonably certain. As with pristane this compound is composed of isoprene units, but limonene is generally considered to be a product of plant biosynthesis. Its origin in this sample is uncertain, but it may arise from plant material ingested by tuna.

Identification and Approximate Quantitation of Major Isolated Organic Constituents

Compound Identification*	Amount Present (ppm)
Heptadiene (?)	0.6
Toluene	0.7
Limonene	0.4
2-nonanone (?)	0.7
2-undecanone (?)	0.1
2,6-di-t-butyl-p-cresol	1.0
Hexadecane	trace
Heptadecane	trace
Pristane	0.03

*Identification followed by a (?) is probable but not definite.

Summary

The analytical values presented in this report represent the authors' evaluation of a considerable amount of data. For many elements where the reports were inconclusive, probable values are not given. If more information becomes available in the future, this report will be updated. Your help in reporting data will be appreciated.

Send any reports to:

Office of Standard Reference Materials (R-701)
B316, Chemistry
National Bureau of Standards
Washington, D.C. 20234

May 12, 1977

U. S. Department of Commerce
Malcolm Baldrige
Secretary

National Bureau of Standards
Ernest Ambler, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1549

Non-Fat Milk Powder

This Standard Reference Material (SRM) is intended primarily for use in calibrating instrumentation and evaluating the reliability of analytical methods for the determination of constituents in milk, milk powders, and other biological matrices.

Certified Values of Constituents: The certified concentrations of the constituent elements are shown in Table 1. Certified values are based on results obtained by definitive methods of known accuracy; or alternatively, from concordant results by two or more independent analytical methods.

Additional Information on Composition: Noncertified concentrations of additional constituent elements are given for information only in Table 2. Noncertified concentrations of lactose and ascorbic acid were determined by high performance liquid chromatography; and for lactose only, by nuclear magnetic resonance.

Notice and Warnings to Users:

Expiration of Certification: This certification is invalid after 3 years from the date of shipping. Should it become invalid before then, purchasers will be notified by NBS.

Stability: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation. The bottle should be kept tightly closed and stored in a desiccator in the dark.

Use: A minimum sample of 500 mg of the dried material (see Instructions for Drying) should be used for any analytical determination to be related to the certified values of this certificate.

Dissolution procedures should be designed to effect complete dissolution, but without losses of volatile elements, such as mercury. Dissolution for these determinations should be carried out in a closed system.

Statistical consultation was provided by K.R. Eberhardt of the Statistical Engineering Division.

The overall direction and coordination of the analyses were under the chairmanship of E.L. Garner, Chief of the Inorganic Analytical Research Division, and W.E. May, Chief of the Organic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Gaithersburg, MD 20899
July 29, 1985
(Revision of Certificates
dated 4-17-84 and
1-14-85)

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

(over)

Instructions for Drying: Samples of this SRM must be dried before weighing according to the following procedure: Dry for 48 hours at 20 to 25 °C in a vacuum oven at a pressure not greater than 30 Pa (0.2 mm Hg).

Analysts:

Center for Analytical Chemistry, National Bureau of Standards:

1. E.S. Beary
2. J.M. Brown-Thomas
3. T.A. Butler
4. B. Coxon
5. M.S. Epstein
6. J.D. Fassett
7. J.W. Gramlich
8. R.R. Greenberg
9. W.R. Kelly
10. H.M. Kingston
11. W.F. Koch
12. G.M. Lambert
13. G.J. Lutz
14. J.R. Moody
15. T.J. Murphy
16. P.J. Paulsen
17. T.C. Rains
18. T.A. Rush
19. M.E. Watson
20. R.L. Watters, Jr.
21. L. Watts

Cooperating Analysts:

22. R.W. Dabeka, Food Research Division, Health Protection Branch, Tunney's Pasture, Ottawa, Ontario, Canada.
23. L. Kosta, A.R. Byrne, M. Dermelj, Institute "Josef Stefan", Ljubljana, Yugoslavia.
24. C. Veillon and K. Patterson, Beltsville Human Nutrition Research Center, U.S. Department of Agriculture, Beltsville, MD.

Table 1. Certified Concentrations of Constituent Elements

Element	Concentration, weight, %	Element	Concentration, weight, %
Calcium[2c,5a]	1.30 ± 0.05	Potassium[2b,5a]	1.69 ± 0.03
Chlorine[3,5a]	1.09 ± .02	Sodium[2c,5a]	0.497 ± .010
Magnesium[2c,5a]	0.120 ± .003	Sulfur[3,4a]	.351 ± .005
Phosphorus[2a,2c]	1.06 ± .02		

Element	Concentration, µg/g	Element	Concentration, µg/g
Cadmium[1b,5b]	0.0005 ± 0.0002	Lead[1b,4a]	0.019 ± 0.003
Chromium[4c,5b]	.0026 ± .0007	Manganese[1b,2a,5a]	.26 ± .06
Copper[1b,2a,5b]	.7 ± .1	Mercury[1a,5b]	.0003 ± .0002
Iodine[4a,6]	3.38 ± .02	Selenium[1d,4b,5a,5b]	11 ± .01
Iron[4a,5a]	1.78 ± .10	Zinc[1c,2c,4b,5a]	46.1 ± 2.2

1. Atomic absorption spectrometry
   a. cold vapor
   b. electrothermal
   c. flame
   d. hydride generation
2. Atomic emission spectrometry
   a. dc plasma
   b. flame
   c. inductively coupled plasma
3. Ion chromatography
4. Isotope dilution mass spectrometry
   a. thermal ionization
   b. spark source
   c. electron impact
5. Neutron activation
   a. instrumental
   b. radiochemical
6. Photon Activation

Notes: (1.) Analytical values are based on the "dry-weight" of material (see Instructions for Drying).

(2.) Except for Fe, the stated uncertainty includes the union of 95% confidence intervals computed separately for each analytical method. It includes the effects of measurement error, possible effects of known systematic errors, and between-method differences. The uncertainty for Fe is given as a 95% confidence interval for the weighted mean of the mass spectrometric and neutron activation values, and includes an allowance (added linearly) for systematic error in the methods. The weights were chosen to minimize the estimated mean squared error of the weighted mean, as described in "Approximately Linear Models," by J. Sacks and D. Ylvisaker, Annals of Statistics 6, pp. 1122-1137, 1978.

Table 2. Noncertified Concentrations of Constituent Elements

Element	Concentration, µg/g	Element	Concentration, µg/g
Aluminum	(2)	Molybdenum	(0.34)
Antimony	(0.00027)	Rubidium	(11)
Arsenic	(.0019)	Silicon	(<50)
Bromine	(12)	Silver	(<0.0003)
Cobalt	(0.0041)	Tin	(<0.02)
Fluorine	(.20)		

Table 3. Noncertified Concentrations of Organic Constituents

Compound	Number of Determinations	Concentration,[a] weight %	Method
Lactose	5	49 ± 3	High Performance Liquid Chromatography
	5	45 ± 2	Proton Nuclear Magnetic Resonance

Compound	Number of Determinations	Concentration,[a] µg/g	Methods
Ascorbic Acid	10	53 ± 5	High Performance Liquid Chromatography

[a]Uncertainties represent one standard deviation.

U. S. Department of Commerce
Malcolm Baldrige
Secretary

National Bureau of Standards
Ernest Ambler, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1566

Oyster Tissue

This Standard Reference Material is intended primarily for use in calibrating instrumentation and validating methodology for the chemical analysis of marine animal tissue.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. Certified values are based on results obtained by reference methods of known accuracy; or alternatively, from results obtained by two or more independent and reliable analytical methods. Non-certified values are given for information only in Table 2. All values are based on a minimum sample size of 250 mg of the dried material.

NOTICE AND WARNINGS TO USERS

Expiration of Certification: This certification is invalid after 5 years from the date of shipping. Should it become invalid before then, purchasers will be notified by NBS.

Storage: The material should be kept tightly closed in its original bottle and stored in a desiccator at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight.

Use: A minimum sample weight of 250 mg of the *dried* material (see Instructions for Drying) is necessary for any certified value in Table 1 to be valid within the stated uncertainty. The bottle should be shaken well before each use, and closed tightly immediately after use.

The statistical analysis of the data was performed by K. R. Eberhardt and H. H. Ku of the Statistical Engineering Division.

The overall direction and coordination of the analytical chemistry measurements leading to this certificate were performed in the NBS Center for Analytical Chemistry by P. D. LaFleur.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
February 22, 1983
(Revision of Certificate
Dated 12-12-79)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

Instructions for Drying: Before weighing, samples of SRM 1566 should be dried to constant weight by one of the following procedures:

1. Reduced-pressure drying at room temperature for 48 hours over $Mg(ClO_4)_2$ in a vacuum desiccator at approximately 1.3×10^4 Pa (100 mm Hg).
2. Vacuum drying at room temperature for 24 hours at a pressure of approximately 30 Pa (0.2 mm Hg) using a cold trap.
3. Freeze drying for 20 hours at a pressure of approximately 3 Pa (0.02 mm Hg).

Source and Preparation of Material: The oysters for this reference material were obtained by the FDA Bureau of Shellfish Sanitation from a commercial source. They had been shucked, frozen, and packaged in sealed plastic bags. The oyster material was ground, freeze-dried, and powdered at the U.S. Army Natick Research and Development Command, Natick, Mass., under the direction of L. Hinnegardt and G. C. Walker. At NBS, preliminary analyses of the material homogeneity indicated that an improvement in homogeneity would be required to establish more reliable certified values for a minimum sample size of 250 mg. Accordingly, the material was cryogenically ground by J. R. Moody and J. Matwey. It was then blended and bottled at NBS, after which it was again freeze-dried at the Natick, Mass., laboratory.

Homogeneity Assessment: Randomly selected bottles of SRM 1566 were sampled and tested for homogeneity by neutron activation and atomic absorption spectrometry. No inhomogeneity was observed for the following elements determined by neutron activation: Na, Cl, V, and Mn. The values for Mg, K, Cu, Zn, and Cd determined by atomic absorption spectrometry were within the imprecision of the method; however, Ca does exhibit some inhomogeneity--approximately 4% relative standard deviation.

Analysts:

Center for Analytical Chemistry, National Bureau of Standards:

1. J. V. Bailey
2. C. Blundell
3. T. J. Brady
4. M. Diaz
5. L. P. Dunstan
6. M. S. Epstein
7. J. D. Fassett
8. M. Gallorini
9. E. L. Garner
10. T. E. Gills
11. J. W. Gramlich
12. R. R. Greenberg
13. S. Hanamura
14. S. Harrison
15. E. F. Heald
16. G. M. Hyde
17. W. R. Kelly
18. H. M. Kingston
19. W. F. Koch
20. R. M. Lindstrom
21. G. J. Lutz
22. L. A. Machlan
23. W. A. MacCrehan
24. E. J. Maienthal
25. J. Maples
26. O. Menis
27. J. D. Messman
28. J. R. Moody
29. L. J. Moore
30. T. J. Murphy
31. P. J. Paulsen
32. T. C. Rains
33. H. L. Rook

Cooperating Analysts:

34. University of Tokyo, Tokyo, Japan; Y. Dokiya (NBS Guest Worker).
35. Division of Chemistry, National Research Council of Canada, Ottawa, Canada; S. Berman, A. Desaulniers, J. McLaren, A. Mykytiuk, D. Russell, and S. Willie.
36. Ibaraki Electrical Communication Laboratory, Nippon Telegraph and Telephone Public Corporation, Tokai, Ibaraki, Japan; K. Kudo and K. Kobayashi.
37. Food Research Division, Health Protection Branch, Tunney's Pasture, Ottawa, Ontario, Canada; R. W. Dabeka, A. D. McKenzie, and H. B. S. Conacher.

Table 1. Certified Values of Constituent Elements

Element[1]	Content[2], Wt. Percent	Element[1]	Content[2], Wt. Percent
Calcium[b,d]	0.15 ± 0.02	Potassium[d]	0.969 ± 0.005
Magnesium[a,d]	0.128 ± 0.009	Sodium[b,f]	0.51 ± 0.03

Element[1]	Content[2], $\mu g/g$	Element[1]	Content[2], $\mu g/g$
Arsenic[a,f,g,h]	13.4 ± 1.9	Nickel[a,e,h]	1.03 ± 0.19
Cadmium[a,d,e,f,h]	3.5 ± 0.4	Rubidium[d,f]	4.45 ± 0.09
Chromium[d,e,f]	0.69 ± 0.27	Selenium[a,e,f]	2.1 ± 0.5
Copper[a,c,e,f]	63.0 ± 3.5	Silver[a,f]	0.89 ± 0.09
Iron[b,c,e,f]	195 ± 34	Strontium[b,d]	10.36 ± 0.56
Lead[a,d,e,h]	0.48 ± 0.04	Uranium[d]	0.116 ± 0.006
Manganese[a,c,f]	17.5 ± 1.2	Vanadium[d]	2.3 ± 0.1
Mercury[a,f]	0.057 ± 0.015	Zinc[a,c,d,e,f,h]	852 ± 14

1. Analytical Methods:
   - [a] Atomic absorption spectroscopy
   - [b] Atomic emission spectroscopy, flame
   - [c] Atomic emission spectroscopy, inductively coupled plasma
   - [d] Isotope dilution mass spectrometry, thermal ionization
   - [e] Isotope dilution mass spectrometry, spark source
   - [f] Neutron activation
   - [g] Photon activation
   - [h] Polarography

2. Based on dry weight. (For drying instructions, see the section of this certificate on Instructions for Drying.) The estimated uncertainty is given as 95 percent tolerance limits for coverage of at least 95 percent of the measured values of all bottles of SRM 1566. For a given element, the following statement can be made at a confidence limit of 95 percent. "If the concentrations were measured for all bottles, at least 95 percent of these measured values should fall within the indicated limits." The concept of tolerance limits is discussed in Chapter 2, Experimental Statistics, NBS Handbook 91, 1966, and page 14, The Role of Standard Reference Materials in Measurement Systems, NBS Monograph 148, 1975.

Table 2. Non-certified Values of Constituent Elements

Element	Content[1] (Wt. Percent)
Chlorine	(1.0)
Sulfur	(0.76)
Phosphorous	(0.81)
	($\mu g/g$)
Bromine	(55)
Cobalt	(0.4)
Fluorine	(5.2)
Iodine	(2.8)
Molybdenum	($\leq$0.2)
Thallium	($\leq$0.005)
Thorium	(0.1)

[1] Based on dry weight. (For drying instructions, see the section of this certificate on Instructions for Drying.)

U.S. Department of Commerce
Juanita M. Kreps
Secretary
National Bureau of Standards
Ernest Ambler, Acting Director

National Bureau of Standards
Certificate of Analysis
Standard Reference Material 1567

Wheat Flour

This Standard Reference Material is intended primarily for calibrating instrumentation and evaluating the reliability of analytical methods for the determination of minor and trace elements in wheat flour and similar agricultural food products.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. They are based on results obtained by two or more independent, reliable analytical methods. Non-certified values which are given for information only, appear in Table 2.

All values are based on a minimum sample size of 400 mg and are reported on a "dry-weight" basis.

Notice and Warnings to Users:

Expiration of Certification: This certification will be invalid after 5 years from the date of shipping. Should it be invalidated before then, purchasers will be notified by NBS.

Storage: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight. Ideally, the bottle should be kept in a desiccator in the dark at the temperature indicated.

Use: The following procedures should be followed to relate the analytical determinations to the values reported in this Certificate. The bottle should be shaken well before each use, and a minimum sample of 400 mg of the material should be used. Selenium and mercury should be determined in the material without drying and the concentration values adjusted for the moisture content of the material using separate samples. Other elements may be determined either on samples without drying as indicated above or on samples vacuum-dried for 24 hr as indicated under "Instructions for Drying."

The overall direction and coordination of the technical measurements leading to this Certificate were performed under the chairmanship of H. L. Rook.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
January 3, 1978

J. Paul Cali, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements[a]

Minor Constituents

Element	Content Wt. Percent
Potassium	0.136 ± 0.004[b]
Calcium	0.019 ± 0.001

Trace Constituents

Element	Content µg/g	Element	Content µg/g
Iron	18.3 ± 1.0	Copper	2.0 ± 0.3
Zinc	10.6 ± 1.0	Selenium	1.1 ± 0.2
Manganese	8.5 ± 0.5	Cadmium	0.032 ± 0.007
Sodium	8.0 ± 1.5	Mercury	0.001 ± 0.0008

[a]Analytical values are based on the "dry-weight" of material (see Instructions for Drying). Selenium and mercury should be determined on samples without drying and the results adjusted to a "dry-weight" basis by determining moisture on separate samples.

[b]The estimated uncertainty is based on judgment and represents an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples 400 mg or more. (No attempt was made to derive exact statistical measures of imprecision because several methods were involved in the determination of the constituents).

Table 2. Non-certified Values for Constituent Elements[a]

NOTE: The values shown in this table are not certified because they are not based on the results of two or more independent reliable methods. These values are included for information only.

Trace Constituents

Element	Content µg/g	Element	Content µg/g
Bromine	(9)	Nickel	(0.18)
Rubidium	(1)	Arsenic	(0.006)
Molybdenum	(0.4)	Tellurium	(≤0.002)

[a]Analytical values are based on the "dry-weight" of material (see Instructions for Drying).

Preparation of Material: The wheat flour for this Standard Reference Material was described by the supplier as milled from a blend of Hard Red Spring and Hard Red Winter wheat grown primarily in South Dakota. The flour was taken from the mill packer during the middle of a run to obtain homogeneous material. The flour had been bleached and brominated in accordance with standard treatments for commercial bakery use. At NBS, the material was passed through a sieve with openings of 425 μm (No. 40) and blended. The bottled material was then subjected to 2.5 megarads of Co-60 radiation for microbiological control at Neutron Products, Inc., Dickerson, Md.

Homogeneity Assessment: A preliminary evaluation of homogeneity was made by instrumental neutron activation using samples of 150 to 300 mg and counting the activities from radionuclides of Mn, K, Zn, Na, and Br. The homogeneity of other certified elements was evaluated using samples of 400 mg or less with the exception of mercury and calcium for which 500 mg and 1 g, respectively, were used. The uncertainties for the concentrations in Table 1 include these results.

Instructions for Drying: Except for selenium and mercury, elements may be determined on samples which have been dried as follows:

Vacuum-dry the material at approximately 25 °C for 24 hours at a pressure not greater than 70 Pa (0.5 mm Hg) with a cold trap at a temperature of about −30 °C or below.

Se and Hg should be determined on undried samples; other elements may be so determined. However, because the Certificate values are reported on a "dry-weight" basis, the elemental concentrations determined on undried samples should be adjusted for the moisture content of the samples. The moisture content, which was approximately 9% when bottled, should be determined on separate samples by either the vacuum-drying procedure described above or drying the sample in air in an oven at 85 °C for 24 hours. Both of these procedures yielded the same loss in weight. Samples for analysis should not be oven-dried lest elements be lost by volatilization.

Analytical Methods Used and Analysts

Analytical Methods

A. Atomic absorption spectrometry
B. Flame emission spectrometry
C. Isotope dilution spark source mass spectrometry
D. Neutron activation
E. Polarography

Analysts

Analytical Chemistry Division, National Bureau of Standards

1. J. R. Baldwin
2. T. J. Brady
3. M. G. Diaz
4. L. P. Dunstan
5. M. S. Epstein
6. M. Gallorini
7. T. E. Gills
8. R. R. Greenberg
9. R. M. Lindstrom
10. G. J. Lutz
11. E. J. Maienthal
12. R. Mavrodineanu
13. J. D. Messman
14. R. M. Morris
15. P. J. Paulsen
16. T. C. Rains
17. P. A. Sleeth

Cooperating Analysts

18. W. R. Wolf and J. Holden, Nutrition Institute, U.S. Department of Agriculture, Beltsville, Md.

U. S. Department of Commerce
Malcolm Baldrige
Secretary
National Bureau of Standards
Ernest Ambler, Director

Addendum to

National Bureau of Standards Certificate of Analysis

Standard Reference Material 1567

Wheat Flour

Additional Certification

The following certified value is to be added to Table 1.

Table 1. Certified Values of Constituent Elements[a]

Element	Content, $\mu g/g$
Lead	0.020 ± 0.010[b]

[a] Lead should be determined on samples without drying and the results adjusted to a "dry-weight" basis by determining moisture on separate samples.
[b] The estimated uncertainty, based on judgment, is for samples 2 g or more.

Analytical Methods Used and Analysts

Inorganic Analytical Research Division, National Bureau of Standards.

Isotope dilution, mass spectrometry, I. L. Barnes and E. S. Beary;
Polarography, E. J. Maienthal.

Cooperating Analyst

R. W. Dabeka, Food Directorate, Health Protection Branch, Ottawa, Canada.

U.S. Department of Commerce
Juanita M. Kreps
Secretary
National Bureau of Standards
Ernest Ambler, Acting Director

National Bureau of Standards
Certificate of Analysis
Standard Reference Material 1568

Rice Flour

This Standard Reference Material is intended primarily for calibrating instrumentation and evaluating the reliability of analytical methods for the determination of minor and trace elements in rice flour and similar agricultural food products.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. They are based on results obtained by two or more independent, reliable analytical methods. Non-certified values which are given for information only, appear in Table 2.

All values are based on a minimum sample size of 400 mg and are reported on a "dry-weight" basis.

Notice and Warnings to Users:

Expiration of Certification: This certification will be invalid after 5 years from the date of shipping. Should it be invalidated before then, purchasers will be notified by NBS.

Storage: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight. Ideally, the bottle should be kept in a desiccator in the dark at the temperature indicated.

Use: The following procedures should be followed to relate the analytical determinations to the values reported in this Certificate. The bottle should be shaken well before each use, and a minimum sample of 400 mg of the material should be used. Selenium and mercury should be determined in the material without drying and the concentration values adjusted for the moisture content of the material using separate samples. Other elements may be determined either on samples without drying as indicated above or on samples vacuum-dried for 24 hr as indicated under "Instructions for Drying."

The overall direction and coordination of the technical measurements leading to this Certificate were performed under the chairmanship of H. L. Rook.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
January 3, 1978

J. Paul Cali, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements[a]

Minor Constituents

Element	Content Wt. Percent
Potassium	0.112 ± 0.002[b]
Calcium	0.014 ± 0.002

Trace Constituents

Element	Content µg/g	Element	Content µg/g
Manganese	20.1 ± 0.4	Arsenic	0.41 ± 0.05
Zinc	19.4 ± 1.0	Selenium	0.4 ± 0.1
Iron	8.7 ± 0.6	Cadmium	0.029 ± 0.004
Sodium	6.0 ± 1.5	Cobalt	0.02 ± 0.01
Copper	2.2 ± 0.3	Mercury	0.0060 ± 0.0007

[a]Analytical values are based on the "dry-weight" of material (see Instructions for Drying). Selenium and mercury should be determined on samples without drying and the results adjusted to a "dry-weight" basis by determining moisture on separate samples.

[b]The estimated uncertainty is based on judgment and represents an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples 400 mg or more. (No attempt was made to derive exact statistical measures of imprecision because several methods were involved in the determination of the constituents).

Table 2. Non-certified Values for Constituent Elements[a]

NOTE: The values shown in this table are not certified because they are not based on the results of two or more independent reliable methods. These values are included for information only.

Trace Constituents

Element	Content µg/g	Element	Content µg/g
Rubidium	(7)	Nickel	(0.16)
Molybdenum	(1.6)	Tellurium	(≤0.002)
Bromine	(1)		

[a]Analytical values are based on the "dry-weight" of material (see Instructions for Drying).

Preparation of Material: The rice flour for this Standard Reference Material was described by the supplier as 100% long grain from Arkansas. At NBS, the material was passed through a sieve with openings of 425 μm (No. 40) and blended. The bottled material was then subjected to 2.5 megarads of Co-60 radiation for microbiological control at Neutron Products, Inc., Dickerson, Md.

Homogeneity Assessment: A preliminary evaluation of homogeneity was made by instrumental neutron activation using samples of 150 to 300 mg and counting the activities from radionuclides of Mn, K, Zn, Na, and Br. The homogeneity of other certified elements was evaluated using samples of 400 mg or less with the exception of mercury and calcium for which 500 mg and 1 g, respectively, were used. The uncertainties for the concentrations in Table 1 include these results.

Instructions for Drying: Except for selenium and mercury, elements may be determined on samples which have been dried as follows:

Vacuum-dry the material at approximately 25 °C for 24 hours at a pressure not greater than 70 Pa (0.5 mm Hg) with a cold trap at a temperature of about −30 °C or below.

Se and Hg should be determined on undried samples; other elements may be so determined. However, because the Certificate values are reported on a "dry-weight" basis, the elemental concentrations determined on undried samples should be adjusted for the moisture content of the samples. The moisture content, which was approximately 9% when bottled, should be determined on separate samples by either the vacuum-drying procedure described above or drying the sample in air in an oven at 85 °C for 24 hours. Both of these procedures yielded the same loss in weight. Samples for analysis should not be oven-dried lest elements be lost by volatilization.

Analytical Methods Used and Analysts

Analytical Methods

A. Atomic absorption spectrometry
B. Flame emission spectrometry
C. Isotope dilution spark source mass spectrometry
D. Neutron activation
E. Polarography

Analysts

Analytical Chemistry Division, National Bureau of Standards

1. J. R. Baldwin
2. T. J. Brady
3. M. G. Diaz
4. L. P. Dunstan
5. M. S. Epstein
6. M. Gallorini
7. T. E. Gills
8. R. R. Greenberg
9. R. M. Lindstrom
10. G. J. Lutz
11. E. J. Maienthal
12. R. Mavrodineanu
13. J. D. Messman
14. R. M. Morris
15. P. J. Paulsen
16. T. C. Rains
17. P. A. Sleeth

Cooperating Analysts

18. W. R. Wolf and J. Holden, Nutrition Institute, U.S. Department of Agriculture, Beltsville, Md

U. S. Department of Commerce
Malcolm Baldrige
Secretary
National Bureau of Standards
Ernest Ambler, Director

Addendum to

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1568

Rice Flour

Additional Certification

The following certified value is to be added to Table 1.

Table 1. Certified Values of Constituent Elements[a]

Element	Concentration, $\mu g/g$
Lead	0.045 ± 0.010[b]

[a] Lead should be determined on samples without drying and the results adjusted to a "dry-weight" basis by determining moisture on separate samples.
[b] The estimated uncertainty, based on judgment, is for samples 2 g or more.

Analytical Methods Used and Analysts

Inorganic Analytical Research Division, National Bureau of Standards.

Isotope dilution, mass spectrometry, I. L. Barnes and E. S. Beary;
Polarography, E. J. Maienthal.

Cooperating Analyst

R. W. Dabeka, Food Directorate, Health Protection Branch, Ottawa, Canada.

U.S. Department of Commerce
Elliot L. Richardson,
Secretary

National Bureau of Standards
Ernest Ambler, Acting Director

National Bureau of Standards
Certificate of Analysis
Standard Reference Material 1569

Brewers Yeast

This Standard Reference Material is intended for use in calibrating instrumentation and evaluating the accuracy of analytical methods for the determination of chromium in brewers yeast and other biological materials. SRM 1569 and like materials contain a volatile chromium component which presents an especially difficult analytical problem. Care should be taken to avoid its loss; see, "Preparation of Biological Materials for Chromium Analysis," W. R. Wolf and F. E. Greene [1].

*Chromium concentration: 2.12 ± 0.05 $\mu g/g$

*Calculated on a dry weight basis from determinations made on samples *without drying*. (See "Precautions" below.) A minimum sample size of 150 mg should be used.

The certified value is based on concordant results by independent analytical methods; the uncertainty is estimated from the imprecision of the methods and inhomogeneity of the material.

The overall direction and coordination of the technical measurements leading to this certificate were performed under the chairmanship of L. McClendon.

The technical aspects leading to the preparation, certification and issuance of this material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

[1] Wolf, W. R. and Greene, F. E., Preparation of Biological Materials for Chromium Analysis, Proceedings of the 7th Materials Research Symposium, Accuracy in Trace Analysis: Sampling, Sample Handling and Analysis, NBS Spec. Publ. 422, U. S. Government Printing Office, Washington, D.C. (August 1976).

Washington, D.C. 20234
September 7, 1976

J. Paul Cali, Chief
Office of Standard Reference Materials

(over)

Analyses were performed in the NBS Analytical Chemistry Division by L. McClendon (neutron activation) and by L. Dunstan and E. Garner (isotope dilution, mass spectrometry). Cooperative analyses were also made by W. R. Wolf, Nutrition Institute, U.S. Department of Agriculture, Beltsville, Md.

The material was furnished by the Nutrition Institute, U.S. Department of Agriculture, Beltsville, Md. At NBS, it was passed through a sieve having openings of 0.15 mm (U.S. Series 100 standard sieve) and blended.

Precautions:

(1) The analytical determinations should be made on samples *without drying*. The determinations should be corrected to a dry weight basis by heating separate samples at 85 °C for 3 hr to determine the weight loss.

(2) Samples should not be dissolved in open vessels.

Material Homogeneity was determined by a neutron activation technique using 150-mg random samples from bottled material representing different locations of the bulk material. The statistical test pattern was proposed by J. Mandel of the NBS Institute for Materials Research.

Stability:

The material should be kept in its original bottle and stored at temperatures between 10-23 °C. Exposure to moisture should be minimized by tightly capping the bottle immediately after use. Ideally, the bottle should be kept in a desiccator at the temperature indicated.

U. S. Department of Commerce
Malcolm Baldrige
Secretary

National Bureau of Standards
Ernest Ambler, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1572

Citrus Leaves

This Standard Reference Material (SRM) is intended primarily for use in calibrating instrumentation and evaluating the reliability of analytical methods for the determination of major, minor, and trace elements in botanical materials, agricultural food products, and similar matrices.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. They are based on results obtained either by definitive methods of known accuracy or by two or more independent analytical methods. Non-certified values, which are given for information only, appear in Table 2.

Notice and Warnings to Users:

Expiration of Certification: This certification is invalid 5 years after the shipping date. Should it be invalidated before then, purchasers will be notified by NBS.

Stability: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation. Ideally, the bottle should be kept tightly closed in a desiccator in the dark at the temperature indicated.

Use: The bottle should be shaken well before each use. A minimum sample of 500 mg of the dried material (see Instructions for Drying) should be used for any analytical determination to be related to the certified values of this certificate.

Statistical consultation was provided by K. Kafadar of the Statistical Engineering Division.

The overall direction and coordination of the analyses leading to this certification were performed under the chairmanship of E.L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
December 20, 1982
(Revision of Certificate
dated 2-22-82)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements

Major and Minor Constituents

Element	Content,[1] (Wt. Percent)
Calcium	3.15 ± 0.10
Magnesium	0.58 ± 0.03
Phosphorus	0.13 ± 0.02
Potassium*	1.82 ± 0.06
Sulfur	0.407 ± 0.009

Trace Constituents

Element	Content,[1] μg/g	Element	Content,[1] μg/g
Aluminum	92 ± 15	Manganese	23 ± 2
Arsenic	3.1 ± 0.3	Mercury	0.08 ± 0.02
Barium	21 ± 3	Molybdenum	0.17 ± 0.09
Cadmium	0.03 ± 0.01	Nickel	0.6 ± 0.3
Chromium	0.8 ± 0.2	Rubidium*	4.84 ± 0.06
Copper	16.5 ± 1.0	Sodium	160 ± 20
Iodine	1.84 ± 0.03	Strontium*	100 ± 2
Iron	90 ± 10	Zinc	29 ± 2
Lead*	13.3 ± 2.4		

[1] Based on dry weight: For drying instructions, see the section of this certificate on Instructions for Drying. The uncertainties are based on judgment and represent an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples weighing 500 mg or more.

*For those elements determined by definitive methods, the uncertainties are given as 95%/95% statistical tolerance intervals. See The Role of Standard Reference Materials in Measurement Systems, NBS Monograph 148, 1975 p 14.

Table 2. Non-certified Values for Constituent Elements

NOTE: The following values are not certified because they are not based on the results of either a definitive method of known accuracy or two or more independent methods. These values are included for information only.

Major Constituent

Element	Content,[1] (Wt. Percent)
Nitrogen	(2.86)

Trace Constituents

Element	Content,[1] µg/g	Element	Content,[1] µg/g
Antimony	(0.04)	Samarium	(0.052)
Bromine	(8.2)	Scandium	(0.01)
Cerium	(0.28)	Selenium	(0.025)
Cesium	(0.098)	Tellurium[a]	(0.02)
Chlorine	(414)	Thallium	(≤0.01)
Cobalt	(0.02)	Tin	(0.24)
Europium	(0.01)	Uranium	(≤0.15)
Lanthanum	(0.19)		

[1]Analytical values are based on the "dry weight" of material (See Instructions for Drying).

[a]Not sufficiently homogeneous for certification.

Instructions for Drying: Samples of this SRM must be dried before weighing and analysis by either of the following procedures:

1. Drying for 2 hours in air in an oven at 85 °C.
2. Drying for 24 hours at 20 to 25 °C and at a pressure not greater than 30 Pa (0.2 mm Hg).

Additional Information on Analyses: This SRM contains siliceous material, which is an integral part of the sample. The values in Tables 1 and 2 are based on analyses performed on the *entire* sample. Therefore, dissolution procedures should be capable of complete dissolution of the sample but should not result in losses of volatile elements, such as arsenic and mercury.

Source and Preparation of Material: The plant material for this SRM was collected and prepared under the direction of A. L. Kenworthy, Michigan State University. Its source was the Lake Alfred area of central Florida. The material was air-dried, ground in a comminuting machine to pass a 425-µm (No. 40) sieve, dried at 85 °C, and thoroughly mixed in a feed blender. After packaging the material in polyethylene-lined fiber drums, it was sterilized in situ with cobalt-60 radiation. The sterilization procedure was carried out at the U.S. Army Research and Development Command, Natick, Mass. under the direction of A. Brynjolfsson.

Analytical Methods Used and Analyses

Analytical Methods:

A. Atomic absorption spectrometry
B. Atomic emission spectrometry, flame
C. Atomic emission spectrometry, inductively coupled plasma
D. Ion chromatography
E. Isotope dilution thermal source mass spectrometry
F. Isotope dilution spark source mass spectrometry
G. Kjeldahl method for nitrogen
H. Neutron activation
I. Photon activation
J. Polarography
K. Spectrophotometry

Analysts:

Inorganic Analytical Research Division, National Bureau of Standards

1. I.L. Barnes
2. E.S. Beary
3. K.A. Brletic
4. T.A. Butler
5. E.R. Deardorff
6. J.W. Gramlich
7. R.R. Greenberg
8. S. Hanamura
9. E.F. Heald
10. W.R. Kelly
11. H.M. Kingston
12. W.F. Koch
13. G.M. Lambert
14. R.M. Lindstrom
15. G.J. Lutz
16. L.A. Machlan
17. E.J. Maienthal
18. J.R. Moody
19. T.J. Murphy
20. P.J. Paulsen
21. L.J. Powell
22. T.C. Rains
23. T.A. Rush
24. P.A. Sleeth
25. R.L. Watters, Jr.
26. R. Zeisler

Cooperating Analysts:

1. M. Ihnat, Chemistry and Biology Research Institute, Agriculture Canada, Ottawa, Canada.
2. M. Gallorini, E. Orvini, and M. DiCasa, Consiglio Nazionale delle Ricerche, Centro di Radiochimica e Analisi per Attivazione presso l'Instituto di Chimica Generale dell' Universitá, Pavia, Italy.
3. L. Kosta, A. R. Byrne, and A. Prosenc, Institute "Josef Stefan," Ljubljana, Yugoslavia.
4. J. B. Jones, Jr., Department of Horticulture, University of Georgia, Athens, Georgia.
5. U. M. Cowgill, Department of Biological Sciences, University of Pittsburgh, Pittsburgh, Pennsylvania.

U.S. Department of Commerce
Elliot L. Richardson,
Secretary

National Bureau of Standards
Ernest Ambler, Acting Director

National Bureau of Standards
Certificate of Analysis
Standard Reference Material 1573
Tomato Leaves

This Standard Reference Material is intended primarily for calibrating instrumentation and evaluating the reliability of analytical methods for the determination of major, minor, and trace elements in botanical materials and other agricultural products.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. They are based on results obtained either by reference methods of known accuracy or by two or more independent, reliable analytical methods. Non-certified values, which are given for information only, appear in Table 2. All values are based on a minimum sample size of 500 mg of the material dried as indicated under "Instructions for Drying."

Notice and Warnings to Users:

Expiration of Certification: This certification will be invalid 5 years after the shipping date. Should it be invalidated before then, purchasers will be notified by NBS.

Stability: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight. Ideally, the bottle should be kept in a desiccator in the dark at the temperature indicated.

Use: The bottle should be shaken well before each use. A minimum sample of 500 mg of the *dried* material (see Instructions for Drying) should be used for any analytical determination to be related to the certified values of this certificate.

The overall direction and coordination of the technical measurements leading to this certificate were performed under the chairmanship of H. L. Rook. The overall coordination of the cooperative work performed by the Commission of European Communities, Joint Research Center, Ispra Establishment, Italy, was by G. Rossi of the Chemistry Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. W. Mears and R. Alvarez

Washington, D.C. 20234
October 18, 1976

J. Paul Cali, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements[a]

Major and Minor Constituents

Element	Content Wt. Percent
Potassium	4.46 ± 0.03
Calcium	3.00 ± 0.03
Phosphorus	0.34 ± 0.02

Trace Constituents

Element	Content μg/g	Element	Content μg/g
Iron	690 ± 25	Lead	6.3 ± 0.3
Manganese	238 ± 7	Chromium	4.5 ± 0.5
Zinc	62 ± 6	Arsenic	0.27 ± 0.05
Strontium	44.9 ± 0.3	Thorium	0.17 ± 0.03
Rubidium	16.5 ± 0.1	Uranium	0.061 ± 0.003
Copper	11 ± 1		

[a]Analytical values are based on the "dry-weight" of material (See Instructions for Drying).

The uncertainties of the values shown in Table 1 include allowances for inhomogeneity, method imprecision, and an estimate of possible biases of the analytical methods used.

Table 2. Non-certified Values for Constituent Elements[a]

NOTE: The following values are not certified because they are not based on the results of either a reference method of known accuracy or two or more independent methods. These values are included for information only.

Major and Minor Constituents

Element	Content Wt. Percent
Nitrogen	(5.0)
Magnesium	(0.7)
Aluminum	(0.12)

Trace Constituents

Element	Content μg/g	Element	Content μg/g
Boron	(30)	Cobalt	(0.6)
Bromine	(26)	Scandium	(0.13)
Cadmium[b]	(3)	Mercury	(0.1)
Cerium	(1.6)	Thallium	(0.05)
Lanthanum	(0.9)	Europium	(0.04)

[a]Analytical values are based on the "dry weight" of material (See Instructions for Drying).

[b]Cadmium was not sufficiently homogeneous for certification.

Instructions for Drying: Samples of this Standard Reference Material *must* be dried before weighing by either of the following procedures:

1. Drying in air in an oven at 85 °C for 2 hours.
2. Lyophilization using a cold trap at or below −50 °C at a pressure *not greater* than 30 Pa (0.2 mm Hg) for 24 hours.

NOTE: Drying either in an oven at 105 °C or in a vacuum oven at 75°C causes large losses of volatiles other than water and should *not* be used.

Additional Information on Analyses: This Standard Reference Material contains siliceous material, which is an integral part of the sample. The analyses reported in Tables 1 and 2 were performed on the entire sample. Therefore, dissolution procedures should be capable of complete dissolution of the sample, but should not result in losses of volatile elements, such as arsenic and mercury.

Source and Preparation of Material: The plant material for this SRM was collected and prepared under the direction of A. L. Kenworthy of Michigan State University, East Lansing, Mich. Its source was a field plot of direct seeded tomatoes that had been established at the Horticultural Research Center of the University. For the preparation of the SRM, the terminal portions of the plants were clipped, air-dried, and ground in a comminuting machine. After grinding, the material was dried at 85 °C, thoroughly mixed in a feed blender, packaged in polyethylene-lined fiber drums, and sterilized in situ with cobalt-60 radiation. The sterilization procedure was carried out at the U.S. Army Research and Development Command, Natick, Mass. under the direction of A. Brynjolfsson. At NBS, a preliminary evaluation of the material homogeneity indicated that its improvement would be required to establish more reliable certified values. Therefore, the material was resieved and the portion that had passed a polypropylene sieve having openings of 0.25 mm (equivalent to a U.S. series 60 standard sieve) was retained for the SRM.

Homogeneity Assessment: Material homogeneity was evaluated by determining nine of the certified elements, P, Fe, Mn, Zn, Rb, Cu, Cr, As, and U on samples of 500 mg or less taken at various locations of the freeze-dried bulk material. The other certified elements, K, Ca, Sr, Pb, and Th were determined using sample weights not exceeding one gram. The uncertainties of the concentrations given in Table 1 include these results.

Analytical Methods Used and Analysts

Analytical Methods

A. Atomic absorption spectroscopy
B. Isotope dilution mass spectrometry
C. Isotope dilution spark source mass spectrometry
D. Kjeldahl method for nitrogen
E. Neutron activation
F. Nuclear track technique
G. Optical emission spectroscopy
H. Spectrophotometry
I. Polarography

(over)

Analysts

Analytical Chemistry Division, National Bureau of Standards

1. R. W. Burke
2. B. S. Carpenter
3. E. R. Deardorff
4. B. I. Diamondstone
5. L. J. Dunstan
6. M. S. Epstein
7. R. H. Filby
8. E. L. Garner
9. T. E. Gills
10. J. W. Gramlich
11. S. H. Harrison
12. R. M. Lindstrom
13. L. A. Machlan
14. L. T. McClendon
15. L. J. Moore
16. T. J. Murphy
17. P. J. Paulsen
18. T. C. Rains
19. H. L. Rook

Cooperating Analysts

20. Chemistry Division, Standards and Reference Substances Secretariat, Commission of European Communities, Joint Research Center, Ispra Establishment, Italy.

    G. Serrini, G. Renaux, W. Leyendecker, E. Orthmann, R. Pietra, G. Guzzi, F. Colombo, F. Girardi, N. Toussaint

21. Y. Nemoto, K. Okamoto, and K. Fuwa, Division of Chemistry and Physics, National Institute for Environmental Studies, Yatabe, Ibaraki, Japan.
22. L. Kosta, Institute "Josef Stefan," Ljubljana, Yugoslavia.
23. J. B. Jones, Jr. and R. Isaac, University of Georgia, Athens, Georgia.

U.S. Department of Commerce
Elliot L. Richardson,
Secretary

National Bureau of Standards
Ernest Ambler, Acting Director

National Bureau of Standards Certificate of Analysis

Standard Reference Material 1575

Pine Needles

This Standard Reference Material is intended primarily for calibrating instrumentation and evaluating the reliability of analytical methods for the determination of major, minor, and trace elements in botanical materials and other agricultural products.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. They are based on results obtained either by reference methods of known accuracy or by two or more independent, reliable analytical methods. Non-certified values, which are given for information only, appear in Table 2. All values are based on a minimum sample size of 500 mg of the material dried as indicated under "Instructions for Drying."

Notice and Warnings to Users:

Expiration of Certification: This certification will be invalid 5 years after the shipping date. Should it be invalidated before then, purchasers will be notified by NBS.

Stability: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight. Ideally, the bottle should be kept in a desiccator in the dark at the temperature indicated.

Use: The bottle should be shaken well before each use. A minimum sample of 500 mg of the *dried* material (see Instructions for Drying) should be used for any analytical determination to be related to the certified values of this certificate.

The overall direction and coordination of the technical measurements leading to this certificate were performed under the chairmanship of H. L. Rook. The overall coordination of the cooperative work performed by the Commission of European Communities, Joint Research Center, Ispra Establishment, Italy, was by G. Rossi of the Chemistry Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. W. Mears and R. Alvarez.

Washington, D.C. 20234
October 18, 1976

J. Paul Cali, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements[a]

Major and Minor Constituents

Element	Content Wt. Percent
Calcium	0.41 ± 0.02
Potassium	0.37 ± 0.02
Phosphorus	0.12 ± 0.02

Trace Constituents

Element	Content µg/g	Element	Content µg/g
Manganese	675 ± 15	Copper	3.0 ± 0.3
Aluminum	545 ± 30	Chromium	2.6 ± 0.2
Iron	200 ± 10	Arsenic	0.21 ± 0.04
Rubidium	11.7 ± 0.1	Mercury	0.15 ± 0.05
Lead	10.8 ± 0.5	Thorium	0.037 ± 0.003
Strontium	4.8 ± 0.2	Uranium	0.020 ± 0.004

[a]Analytical values are based on the "dry-weight" of material (See Instructions for Drying).

The uncertainties of the values of the constituents shown in Table 1 include allowances for material inhomogeneity, method imprecision, and an estimate of possible biases of the analytical methods used.

Table 2. Non-certified Values for Constituent Elements[a]

NOTE: The following values are not certified because they are not based on the results of either a reference method of known accuracy or two or more independent methods. These values are included for information only.

Major Constituent

Element	Content Wt. Percent
Nitrogen	(1.2)

Trace Constituents

Element	Content µg/g	Element	Content µg/g
Bromine	(9)	Lanthanum	(0.2)
Nickel	(3.5)	Cobalt	(0.1)
Cerium	(0.4)	Thallium	(0.05)
Cadmium[b]	(<0.5)	Scandium	(0.03)
Antimony	(0.2)	Europium	(0.006)

[a]Analytical values are based on the "dry-weight" of material (See Instructions for Drying).

[b]Cadmium was not sufficiently homogeneous for certification.

Instructions for Drying: Samples of this Standard Reference Material *must* be dried before weighing by either of the following procedures:

1. Drying in air in an oven at 85 °C for 2 hours.
2. Lyophilization using a cold trap at or below −50 °C at a pressure *not greater* than 30 Pa (0.2 mm Hg) for 24 hours.

NOTE: Drying either in an oven at 105 °C or in a vacuum oven at 75°C causes large losses of volatiles other than water and should *not* be used.

Additional Information on Analyses: This Standard Reference Material contains siliceous material, which is an integral part of the sample. The analyses reported in Tables 1 and 2 were performed on the entire sample. Therefore, dissolution procedures should be capable of complete dissolution of the sample but should not result in losses of volatile elements, such as arsenic and mercury.

Source and Preparation of Material: The plant material for this SRM was collected and prepared under the direction of A. L. Kenworthy of Michigan State University, East Lansing, Mich. Its source was Manistee State Park, approximately 65 km north of Muskegon, Mich. For the preparation of the SRM, the material was air-dried, and ground in a comminuting machine. After grinding the material, it was dried at 85 °C, thoroughly mixed in a feed blender, packaged in polyethylene-lined fiber drums, and sterilized in situ with cobalt-60 radiation. The sterilization procedure was carried out at the U.S. Army Research and Development Command, Natick, Mass. under the direction of A. Brynjolfsson. At NBS, preliminary evaluation of the material homogeneity indicated that its improvement would be required to establish more reliable certified values. Therefore, the material was resieved and the portion that passed a polypropylene sieve having openings of 0.25 mm (equivalent to a U.S. series 60 standard sieve) was retained for the SRM.

Homogeneity Assessment: Material homogeneity was evaluated by determining ten of the certified elements, P, Al, Fe, Mn, Rb, Cu, Cr, As, Hg, and U on samples of 500 mg or less taken at various locations of the freeze-dried bulk material. The other certified elements, K, Ca, Sr, Pb, and Th were determined using sample weights not exceeding one gram. The uncertainties for the concentrations given in Table 1 include these results.

Analytical Methods Used and Analysts

Analytical Methods

A. Atomic absorption spectroscopy
B. Isotope dilution mass spectrometry
C. Isotope dilution spark source mass spectrometry
D. Kjeldahl method for nitrogen
E. Neutron activation
F. Nuclear track technique
G. Optical emission spectroscopy
H. Spectrophotometry
I. Polarography

(over)

Analysts

Analytical Chemistry Division, National Bureau of Standards

1. R. W. Burke
2. B. S. Carpenter
3. E. R. Deardorff
4. B. I. Diamondstone
5. L. J. Dunstan
6. M. S. Epstein
7. R. H. Filby
8. E. L. Garner
9. T. E. Gills
10. J. W. Gramlich
11. S. H. Harrison
12. R. M. Lindstrom
13. L. A. Machlan
14. E. J. Maienthal
15. L. T. McClendon
16. L. J. Moore
17. T. J. Murphy
18. P. J. Paulsen
19. H. L. Rook

Cooperating Analysts

20. Chemistry Division, Standards and Reference Substances Secretariat, Commission of European Communities, Joint Research Center, Ispra Establishment, Italy.

    G. Guzzi
    A. Colombo
    F. Girardi
    R. Pietra
    N. Toussaint

21. Y. Nemoto, K. Okamoto, and K.Fuwa, Division of Chemistry and Physics, National Institute for Environmental Studies, Yatabe, Ibaraki, Japan.
22. R. Schelenz, Federal Research Center for Nutrition, Karlsruhe, West Germany.
23. L. Kosta, Institute "Josef Stefan," Ljubljana, Yugoslavia.
24. J. B. Jones, Jr. and R. Isaac, University of Georgia, Athens, Georgia.

U. S. Department of Commerce
Malcolm Baldrige
Secretary

National Bureau of Standards
Ernest Ambler, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1577a

Bovine Liver

This Standard Reference Material (SRM) is intended primarily for use in calibrating instrumentation and evaluating the reliability of analytical methods for the determination of major, minor, and trace elements in animal tissue and other biological matrices.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. Certified values are based on results obtained by definitive methods of known accuracy; or alternatively, from results obtained by two or more independent analytical methods. Noncertified values are given for information only in Table 2.

Notice and Warnings to Users:

Expiration of Certification: This certification is invalid after 5 years from the date of shipping. Should it become invalid before then, purchasers will be notified by NBS.

Stability: The material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense sources of radiation. The bottle should be kept tightly closed and stored in a desiccator in the dark.

Use: A minimum sample of 250 mg of the dried material (see Instructions for Drying) should be used for any analytical determination to be related to the certified values of this Certificate.

Dissolution procedures should be designed to effect complete solution, but without losses of volatile elements, such as mercury. Dissolution for these determinations should be carried out in a closed system.

Statistical consultation was provided by K.R. Eberhardt and T.R. Crichton of the Statistical Engineering Division.

The overall direction and coordination of the analyses leading to this certification were under the chairmanship of E.L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Gaithersburg, MD 20899
February 1, 1985
(Revision of Certificates
dated 3-5-82, 6-15-82)

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements

Element	Content,[a] (Wt. Percent)
Chlorine	0.28 ± 0.01
Phosphorus	1.11 ± 0.04
Potassium*	0.996 ± 0.007
Sodium	0.243 ± 0.013
Sulfur	0.78 ± 0.01

Element	Content,[a] (μg/g)	Element	Content,[a] (μg/g)
Arsenic	0.047 ± 0.006	Mercury	0.004 ± 0.002
Cadmium	0.44 ± 0.06	Molybdenum	3.5 ± 0.5
Calcium	120 ± 7	Rubidium*	12.5 ± 0.1
Cobalt	0.21 ± 0.05	Selenium	0.71 ± 0.07
Copper	158 ± 7	Silver	0.04 ± 0.01
Iron	194 ± 20	Strontium*	0.138 ± 0.003
Lead*	0.135 ± 0.015	Uranium*	0.00071 ± 0.00003
Magnesium	600 ± 15	Vanadium*	0.099 ± 0.008
Manganese	9.9 ± 0.8	Zinc	123 ± 8

[a]Dry weight: For drying instructions, see the section of this Certificate on Instructions for Drying.

The estimated uncertainties are based on judgment and represent an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples weighing 250 mg or more.

*For those elements determined by definitive methods, the uncertainties are given as 95%/95% statistical tolerance limits. See "The Role of Standard Reference Materials in Measurement Systems," NBS Monograph 148, 1975 p 14.

Table 2. Noncertified Values of Constituent Elements

Element	Content,[a] (Wt. Percent)
Nitrogen	(10.7)

Element	Content,[a] μg/g
Aluminum	(2)
Antimony	(0.003)
Bromine	(9)
Thallium	(0.003)

[a]Dry weight: For drying instructions, see the section of this Certificate on Instructions for Drying.

Instructions for Drying: Samples of this SRM must be dried before weighing according to the following procedure: Dry for 24 hours at 20 to 25 °C in a vacuum oven at a pressure not greater than 30 Pa (0.2 mm Hg).

Source and Preparation of Material:

The liver for this standard was obtained in the Portland, Oregon, area. The gross fat, major blood vessels, and "skin" were removed and the liver was ground. The ground liver was then mixed, transferred to polyethylene-lined trays, and lyophilized by Oregon Freeze Dry Foods, Inc., Albany, Oregon. After lyophilization, the liver was powdered in a Tornado mill, packaged in moisture-proof bags, and then transported to the National Bureau of Standards.

Analysts and Analytical Methods Used

Analytical Methods:

A. Atomic absorption spectrometry
B. Atomic emission spectrometry, flame
C. Atomic emission spectrometry, inductively coupled plasma
D. Ion chromatography
E. Isotope dilution thermal source mass spectrometry
F. Isotope dilution spark source mass spectrometry
G. Kjeldahl method for nitrogen
H. Neutron activation
I. Polarography
J. Spectrophotometry

Analysts:

Analytical Chemistry Division, National Bureau of Standards:

1. J.V. Bailey
2. I.L. Barnes
3. E.S. Beary
4. C.G. Blundell
5. K.A. Brletic
6. T.A. Butler
7. E.R. Deardorff
8. M.S. Epstein
9. J.D. Fassett
10. J.W. Gramlich
11. R.R. Greenberg
12. S. Hanamura
13. S.H. Harrison
14. E.F. Heald
15. W.R. Kelly
16. H.M. Kingston
17. W.F. Koch
18. G.M. Lambert
19. R.M. Lindstrom
20. G.J. Lutz
21. L.A. Machlan
22. E.J. Maienthal
23. T.J. Murphy
24. P.J. Paulsen
25. L.J. Powell
26. T.C. Rains
27. T.A. Rush
28. R.L. Watters, Jr.
29. R. Zeisler

Cooperating Analysts:

30. M. Gallorini and E. Orvini, Consiglio Nazionale delle Ricerche, Centro di Radiochimica e Analisi per Attivazione presso l' Instituto di Chimica Generale dell'Universita, Pavia, Italy.

31. L. Kosta, A.R. Byrne, M. Dermelj, Institute "Josef Stefan", Ljubljana, Yugoslavia.

NBS Technical Publications

Periodical

Journal of Research—The Journal of Research of the National Bureau of Standards reports NBS research and development in those disciplines of the physical and engineering sciences in which the Bureau is active. These include physics, chemistry, engineering, mathematics, and computer sciences. Papers cover a broad range of subjects, with major emphasis on measurement methodology and the basic technology underlying standardization. Also included from time to time are survey articles on topics closely related to the Bureau's technical and scientific programs. Issued six times a year.

Nonperiodicals

Monographs—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

Handbooks—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

Special Publications—Include proceedings of conferences sponsored by NBS, NBS annual reports, and other special publications appropriate to this grouping such as wall charts, pocket cards, and bibliographies.

Applied Mathematics Series—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

National Standard Reference Data Series—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a worldwide program coordinated by NBS under the authority of the National Standard Data Act (Public Law 90-396). NOTE: The Journal of Physical and Chemical Reference Data (JPCRD) is published quarterly for NBS by the American Chemical Society (ACS) and the American Institute of Physics (AIP). Subscriptions, reprints, and supplements are available from ACS, 1155 Sixteenth St., NW, Washington, DC 20056.

Building Science Series—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

Technical Notes—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

Voluntary Product Standards—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The standards establish nationally recognized requirements for products, and provide all concerned interests with a basis for common understanding of the characteristics of the products. NBS administers this program as a supplement to the activities of the private sector standardizing organizations.

Consumer Information Series—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

Order the **above** *NBS publications from: Superintendent of Documents, Government Printing Office, Washington, DC 20402.*

Order the **following** *NBS publications—FIPS and NBSIR's—from the National Technical Information Service, Springfield, VA 22161.*

Federal Information Processing Standards Publications (FIPS PUB)—Publications in this series collectively constitute the Federal Information Processing Standards Register. The Register serves as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations).

NBS Interagency Reports (NBSIR)—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Service, Springfield, VA 22161, in paper copy or microfiche form.

U.S. Department of Commerce
National Bureau of Standards
Gaithersburg, MD 20899

Official Business
Penalty for Private Use $300